


Dash No.	C
-01	.0080
-02	.0050
-03	.0025
-04	.0010
-05	.0100

NOTES:

1. PRIOR TO GRINDING, ENGRAVE OR STAMP SERIAL NUMBER ON THE SIDE WITH FOLLOWING FORMAT: AWWYYSS, THE LETTER 'A', TWO DIGIT WEEK NUMBER, TWO DIGIT YEAR NUMBER, SEQUENCE NUMBER STARTING WITH '00' REPRESENTING THE NUMBER OF CALIBRATION BLOCKS MADE THAT WEEK.
2. APPLY LABEL SHOWING SERIAL NUMBER, DATE OF CALIBRATION, EXPIRATION DATE.
3. CAL. LAB TO PROVIDE BLOCK LABEL INFORMATION, LAB CONDITIONS, UNCERTAINTY, EQUIPMENT USED, DATE CALIBRATED AND EXPIRATION, S/N.
4. FIRST TIME CALIBRATIONS SHOULD HAVE EXPIRATION DATE BLANK, RECERTIFICATIONS SHOULD SHOW ONE YEAR EXPIRATION DATES.
5. SURFACE -A-, LOWER SURFACE, AND SURFACE -B-, UPPER SURFACE, SHALL BE MEASURED USING 40 DATA POINTS EACH ON A 0.2 INCH GRID, AS SHOWN. THE FLATNESS, PARALLELISM, AND RELATIVE HEIGHT OF REGIONS SHALL BE REPORTED ON THE CALIBRATION SHEET. FLATNESS SHALL BE MEASURED FROM THE DIFFERENCE IN HEIGHT AROUND A BEST FIT PLANE (BFP). PARALLELISM SHALL BE MEASURED BY TRANSLOCATING A PARALLEL SURFACE -A- OVER SURFACE -B-, THEN MEASURING THE DIFFERENCE IN HEIGHT SURFACE -B- MAKES TO THE TRANSLOCATED SURFACE. THE DIFFERENCE IN HEIGHT SHALL BE MEASURED BY SUBTRACTING THE GEOMETRIC CENTROIDS BETWEEN THE TWO SURFACES ALONG A VECTOR PERPENDICULAR TO SURFACE -A-. THE DATA SHOULD BE REPORTED TO FIVE DECIMAL PLACES, WHERE EQUIPMENT AND METHODS PERMIT.
6. ENGRAVE OR STAMP PART NUMBER ON SIDE SHOWN IN THE FOLLOWING FORMAT: 30-20MP-0112-XX ('XX' BEING THE DASH NO. PER 'HEIGHT DIFFERENCE TABLE' SHOWN).



PROPRIETARY AND CONFIDENTIAL THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF AKROMETRIX, LLC. ANY REPRODUCTION IN PART OR AS A WHOLE WITHOUT THE WRITTEN PERMISSION OF AKROMETRIX, LLC IS PROHIBITED.		UNLESS OTHERWISE SPECIFIED: DIMENSIONS ARE IN INCHES DO NOT SCALE FRACTIONAL: $\pm 1/32"$ ANGULAR: $\pm .03^\circ$ HOLE: $\pm .015"$ 2 PLACE DECIMAL: $\pm .005"$ 3 PLACE DECIMAL: $\pm .0005"$		 akrometrix 2700 Northeast Expressway, Suite B 500, Atlanta, GA 30345	
MATERIAL: 17-4 PH STAINLESS STEEL 0.375 THICK		TITLE: Calibration Block Shadow Moiré			
FINISH: GLASS BEAD BLAST 175-325 MESH		SIZE B	DRAWING NO. 30-20MP-0112	SHEET 1 OF 1	REV. E