

**SPI to 1" Range, 0.0005" Graduation, Electronic Drop Indicator
Flat & Lug Back, Accurate to 0.0012", Inch & Metric System, Digital
Display, NIST Traceability Certification**

Specs

Minimum Measurement (Decimal Inch)	0.0000
Minimum Measurement (Inch)	0
Minimum Measurement (mm)	0.00
Maximum Measurement (Inch)	1
Maximum Measurement (mm)	25.00
Maximum Measurement (Decimal Inch)	1.0000
Resolution (Decimal Inch)	0.000500
Resolution (mm)	0.0100
Back Type	Flat and Lug
Unit of Measure	Inch; Metric
Display Type	Digital
Traceability Certification Included	Nist
Accuracy (Decimal Inch)	0.001200
Accuracy (mm)	0.0200
Measuring Force (N)	1.50