

Manual Supplement

Manual Title:	123B/124B/125B Calibration	Supplement Issue:	4
Part Number:	Web-Only	Issue Date:	10/22
Print Date:	May 2016	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual..

Change #1, 347

On page 40 , Table 17, change the **Input Value** for Cap Gain (CL 0566):

From: 50 nF

To: 500 nF

Change #2, 387, 613

On page 22, replace Step 6c with:

1. Noise filter: **OFF**

On page 23, replace Step 6g with:

2. Noise filter: **OFF**

Change #3, a272, 676

On page 72, under Input A and Input B, for DC Voltage (VDC) replace Accuracy with:

Accuracy..... $\pm(0.5 \% + 5 \text{ counts})$ with noise filter OFF, no signal applied, after AUTOSSET, and the ground level moved to the center grid line.