



Agilent Technologies 85025A/B/D/E Detectors

User's and Service Guide

REPRODUCTION AND DISTRIBUTION OF
THIS TECHNICAL MANUAL IS AUTHORIZED
FOR GOVERNMENT PURPOSES.



AgilentTechnologies

User's and Service Guide

This User's and Service Guide provides instructions on installing, operation, and maintenance of the Agilent Technologies Model 85025A/B/D/E Detectors supplied with the Agilent Technologies Model 8757D-E02 Scalar Network Analyzer being delivered under U.S. Navy Contract Number N00104-07-D-D014.

Manual Part Number 85025-90073

November 2007

REPRODUCTION AND DISTRIBUTION OF THIS
TECHNICAL MANUAL IS AUTHORIZED FOR
GOVERNMENT PURPOSES.

© Copyright Agilent Technologies 2007
All Rights Reserved

AGILENT TECHNOLOGIES INC
DIVISION TMO CUSTOMER BUSINESS CENTER
9780 S MERIDIAN BLVD
ENGLEWOOD CO 80112 – 5910

Agilent Technologies
85025A/B/D/E Detectors

Hewlett-Packard to Agilent Technologies Transition

This documentation supports a product that previously shipped under the Hewlett-Packard company brand name. The brand name has now been changed to Agilent Technologies. The two products are functionally identical, only our name has changed. The document still includes references to Hewlett-Packard products, some of which have been transitioned to Agilent Technologies.

Certification

Agilent Technologies certifies that this product met its published specifications at the time of shipment from the factory. Agilent further certifies that its calibration measurements are traceable to the United States National Institute of Standards and Technology, to the extent allowed by the Institute's calibration facility, and to the calibration facilities of other International Standards Organization members.

DOCUMENTATION WARRANTY

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, AGILENT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. AGILENT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD AGILENT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CONFLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT WILL CONTROL.

EXCLUSIVE REMEDIES

THE REMEDIES PROVIDED HEREIN ARE BUYER'S SOLE AND EXCLUSIVE REMEDIES. AGILENT SHALL NOT BE LIABLE FOR ANY DIRECT, INDIRECT, SPECIAL, INCIDENTAL, OR CONSEQUENTIAL DAMAGES, WHETHER BASED ON CONTRACT, TORT, OR ANY OTHER LEGAL THEORY.

Assistance

Product maintenance agreements and other customer assistance agreements are available for Agilent products. By internet, phone, or fax, get assistance with all your test & measurement needs.

Contacting Agilent

Online Assistance: www.agilent.com/find/assist

United States

(tel) 1 800 452 4844

Japan

(tel) (81) 426 56 7832

(fax) (81) 426 56 7840

New Zealand

(tel) 0 800 738 378

(fax) 64 4 495 8950

Canada

(tel) 1 877 894 4414

(fax) (905) 206 4120

Latin America

(tel) (305) 269 7500

(fax) (305) 269 7599

Asia Pacific

(tel) (852) 3197 7777

(fax) (852) 2506 9284

Europe

(tel) (31 20) 547 2323

(fax) (31 20) 547 2390

Australia

(tel) 1 800 629 485

(fax) (61 3) 9210 5947

Safety Notes

The following safety notes are used throughout this manual. Familiarize yourself with each of the notes and its meaning before operating this instrument.

Caution Caution denotes a hazard. It calls attention to a procedure that, if not correctly performed or adhered to, would result in damage to or destruction of the instrument. Do not proceed beyond a caution sign until the indicated conditions are fully understood and met.



The instruction documentation symbol. The product is marked with this symbol when it is necessary for the user to refer to the instructions in the documentation.

How to Use This Guide

This guide uses the following conventions:

- Front-Panel Key** This represents a key physically located on the instrument.
- Softkey** This indicates a “softkey,” a key whose label is determined by the instrument’s firmware.
- Screen Text** This indicates text displayed on the instrument’s screen.

Documentation Description

This manual contains information on operating, testing, and servicing the Agilent 85025A/B/D/E detectors.

— |

| —

— |

| —

Contents

1. General Information	
Product Description	1-1
Specifications and Supplemental Characteristics	1-3
Operating Environment	1-9
Accessories	1-9
Storage and Shipment	1-9
Environment	1-9
Packaging	1-10
Returning a Detector for Service	1-10
2. Installation	
Initial Inspection	2-1
Electrostatic Discharge (ESD)	2-2
Static-Safe Work Station	2-2
Static-Safe Practices	2-3
Power Requirements	2-3
Mating Connectors	2-3
Connecting the Detector	2-3
3. Operation	
Features	3-2
Operating Theory	3-2
Measurement System Configuration	3-3
DC Detection mode	3-3
Accurate DC Measurements	3-3
Zeroing the Detector	3-3
Autozero	3-3
Manual Zero	3-4
AC Detection Measurements	3-6
Operator's Check	3-6
Procedure	3-7
If the Operator's Check Fails	3-8

4. Performance Tests	
Equipment Required	4-1
Return Loss Performance Test	4-3
85025A/B/D/E Return Loss Performance Test Procedure	4-3
Specifications	4-3
Description	4-3
Return Loss Measurement	4-6
Calibrating the Scalar Network Analyzer	4-7
Return Loss from 40 MHz to 18 GHz (to 26.5 GHz for 85025B/E only, to 50 GHz for 85025D only)	4-7
If This Test Fails	4-9
Frequency Response Performance Test	4-9
Description	4-9
Equipment Required	4-11
Equipment Common to 85025A/B/D/E	4-11
Additional Test Equipment Required for 85025A	4-11
Additional Test Equipment Required for 85025A Option 001	4-11
Additional Test Equipment Required for 85025B	4-11
Additional Test Equipment Required for 85025D Only	4-11
Additional Test Equipment Required for 85025E Only	4-12
Specifications	4-13
85025A/B/D/E Frequency Response Performance Test Procedure	4-13
Configuring the System	4-13
Characterizing the Source	4-14
Characterizing the Detector	4-14
Computing the Maximum Error	4-14
Power Accuracy Performance Test	4-15
Specifications	4-15
Description	4-15
Equipment Required	4-16
Procedure	4-17
Absolute Power Accuracy in DC Mode Performance Test	4-17
Dynamic Accuracy in AC Mode Performance Test	4-20
Power Accuracy, Alternate Procedure Using an 8350B (+10 dBm maximum)	4-21
Alternate Equipment	4-21
Procedure	4-22
Absolute Power Accuracy in DC Mode, Alternate Procedure	4-23
Dynamic Accuracy in AC Mode, Alternate Procedure	4-24
Performance Test Record	4-25

5. Adjustments	
6. Replaceable Parts	
How To Order Parts Fast	6-1
7. Service	
Repair	7-2
Removing the Covers	7-2
Procedure	7-2
Replacing the Detector	7-3
Replacing the Cable Assembly	7-3
Replacing the Connectors	7-5
8. Detector Maintenance	
Mechanical Inspection	8-1
Inspecting the Connectors	8-1
Visual Examination	8-2
Cleaning the Connectors	8-2
Connector Cleaning Kit	8-2
Gaging Connectors	8-2
Gaging Connectors to be Mated with the 85025A/B/D/E	8-3
Type-N female	8-3
Precision 7 mm	8-3
Precision 3.5 mm female	8-3
9. Automated Program Listing	
Automating the Frequency Response Test	9-1
Cal Factor Entry Program	9-2
Running the Cal Factor Entry Program	9-2
Detector Frequency Response Program	9-3
Running the Detector Frequency Response Program	9-3
Measurement Setup	9-5
Example Programs	9-5

Index

Figures

1-1. 85025A/B/D/E Detector	1-2
2-1. Example of a Static-Safe Work Station	2-2
3-1. Detector Features	3-2
3-2. Typical System Setup for 0.01 to 50 GHz Measurements	3-5
3-3. Operator's Check Equipment Setup	3-7
4-1. Setup for 85025A/B/D/E Return Loss Test	4-6
4-2. 85025A/B/E Return Loss 0.04 GHz to Maximum Frequency	4-8
4-3. Frequency Response Measurement Setup	4-12
4-4. Absolute Power Accuracy Test Setup	4-17
4-5. Power Accuracy Alternate Test Setup	4-22
6-1. Module Exchange Program	6-3
7-1. Removing the Detector Covers	7-2
7-2. Cable Connections	7-5
9-1. Typical Program Output	9-4

Tables

1-1. 85025 Series Detector Descriptions	1-1
1-2. 85025A/B/D/E Detector General Specifications	1-3
1-3. 85025A Detector Specifications (including Option 001)	1-4
1-4. 85025B Detector Specifications	1-5
1-5. 85025D Detector Specifications	1-6
1-6. 85025E Detector Specifications	1-7
1-7. 85025A/B/D/E Detector Supplemental Characteristics	1-8
3-1. Equipment Required for Operator's Check	3-6
4-1. Recommended Equipment	4-2
4-2. 85025A/B Return Loss with Measurement Uncertainty	4-5
4-3. 85025D Return Loss with Measurement Uncertainty	4-5
4-4. 85025E Return Loss with Measurement Uncertainty	4-5
4-5. Approximate Error Analysis at 18 GHz for 85025A/B/E Detectors	4-10
4-6. Approximate Error Analysis for the 85025D Detector	4-10
6-1. 85025A/B/D/E Replaceable Parts and Accessories	6-2

— |

| —

— |

| —

General Information


This manual contains information on operating, testing, and servicing the Agilent 85025A/B/D/E detectors. Figure 1-1 shows the detectors.

Product Description

The 85025A/B/D/E detectors are specifically designed for use with Agilent 8757 series scalar network analyzers.

Table 1-1. 85025 Series Detector Descriptions

Detector	Connector Type	Frequency Range
85025A	Type-N (m)	.01 to 18 GHz
85025A Option 001	precision 7 mm	.01 to 18 GHz
85025B	precision 3.5 mm (m)	.01 to 26.5 GHz
85025D	precision 2.4 mm (m)	.01 to 50 GHz
85025E	precision 3.5 mm (m)	.01 to 26.5 GHz



pc41a_e

Figure 1-1. 85025A/B/D/E Detector

Specifications and Supplemental Characteristics

Tables 1-2 through 1-6 list detector specifications when used with an 8757 series scalar network analyzer. These specifications represent the warranted performance standards or limits against which you can test the device.

Table 1-7 lists supplemental (typical, non-warranted) detector characteristics, when used with one of the above-mentioned analyzers.

Note Specifications describe the instrument's warranted performance over the temperature range of 25 °C, ± 5 °C.

Table 1-2. 85025A/B/D/E Detector General Specifications

Dynamic Range (on all 8757 Series analyzer's detector inputs):	
AC mode	+16 to -55 dBm
DC mode	+16 to -50 dBm
Nominal Impedance	50 Ω
Maximum Input Power	+20 dBm (100 mW), ± 10 VDC

**Table 1-3.
85025A Detector Specifications (including Option 001)**

Frequency Range	0.01 to 18 GHz
Return Loss:	
10 MHz to 40 MHz	10 dB
40 MHz to 4 GHz	20 dB
4 GHz to 18 GHz	17 dB
Frequency Response (in DC mode, input power - 10 dBm):	
10 MHz to 40 MHz	+ 0.25 dB/-0.75 dB
40 MHz to 18 GHz	± 0.5 dB
Absolute Power Accuracy (in DC mode, 50 MHz, calibrated at 0 dBm)	
<p align="right">pc42a_e</p>	
Dynamic Power Accuracy	
<p align="center">AC Mode 50 MHz Calibrated at 0dBm</p> <p align="right">pc43a_e</p>	

Table 1-4. 85025B Detector Specifications

Frequency Range	0.01 to 26.5 GHz
Return Loss:	
10 MHz to 40 MHz	10 dB
40 MHz to 4 GHz	20 dB
4 GHz to 18 GHz	17 dB
18 GHz to 26.5 GHz	12 dB
Frequency Response (in DC mode, input power – 10 dBm):	
10 MHz to 40 MHz	+0.25 dB/-0.75 dB
40 MHz to 18 GHz	±0.5 dB
Absolute Power Accuracy (in DC mode, 50 MHz, calibrated at 0 dBm)	
<p>pc44a_e</p>	
Dynamic Power Accuracy	
<p>AC Mode 50 MHz Calibrated at 0dBm</p> <p>pc45a_e</p>	

Table 1-5. 85025D Detector Specifications

Frequency Range	0.01 to 50 GHz
Return Loss:	
10 MHz to 40 MHz	10 dB
40 MHz to 100 MHz	20 dB
100 MHz to 14 GHz	23 dB
14 GHz to 34 GHz	20 dB
34 GHz to 40 GHz	15 dB
40 GHz to 50 GHz	9 dB
Frequency Response (in DC mode, input power –10 dBm):	
10 MHz to 40 MHz	+ 0.25 dB/–0.75 dB
40 MHz to 20 GHz	±0.5 dB
20 GHz to 26.5 GHz	+ 1/–0.5 dB
26.5 GHz to 40 GHz	+ 2.5/–0.5 dB
40 GHz to 50 GHz	+ 3.0/–0.5 dB
Absolute Power Accuracy (in DC mode, 50 MHz, calibrated at 0 dBm)	
<p style="text-align: right;">pc46a_e</p>	
Dynamic Power Accuracy	
<p style="text-align: center;">AC Mode Mode 50 MHz Calibrated at 0dBm</p> <p style="text-align: right;">pc47a_e</p>	

Table 1-6. 85025E Detector Specifications

Frequency Range	0.01 to 26.5 GHz
Return Loss:	
10 MHz to 40 MHz	10 dB
40 MHz to 100 MHz	20 dB
100 MHz to 25 GHz	25 dB
25 GHz to 26.5 GHz	23 dB
Frequency Response (in DC mode, input power – 10 dBm):	
10 MHz to 40 MHz	+0.25 dB/-0.75 dB
40 MHz to 18 GHz	±0.5 dB
18 GHz to 26.5 GHz	±0.5 dB at 18 GHz to ±1.4 dB at 26.5 GHz
Absolute Power Accuracy (in DC mode, 50 MHz, calibrated at 0 dBm)	
<p>Max Error (+/- dB)</p> <p>Power (dBm)</p> <p>pc44a_e</p>	
Dynamic Power Accuracy	
<p>AC Mode 50 MHz Calibrated at 0dBm</p> <p>Max Error (+/- dB)</p> <p>Change in Power (dB)</p> <p>pc45a_e</p>	

**Table 1-7.
85025A/B/D/E Detector Supplemental Characteristics**

RF Connector Mechanical Tolerances:	
Recession of the male center conductor from reference plane:	
85025A	0.207 to 0.210 inches ¹
85025A Option 001	0.000 to 0.003 inches
85025B	0.000 to 0.003 inches
85025D	0.000 to 0.002 inches
85025E	0.000 to 0.003 inches
Cable Length	1.22 m (48 inches)
Weight (Including cable):	
Net:	0.24 kg (0.5 lb)
Shipping:	1.0 kg (2.2 lb)
Dimensions² (Including input connector, not including cable)	

1 Because a type-N gage calibration block zeros the gage at a 0.207-inch offset, the gage displays a 0.207- to 0.210-inch offset as 0.000 to 0.003 inches.

2 The model used in this illustration is an 85025A. Because of varying input connector lengths, the overall length measurements for the other detector models covered by this manual are:

- 85025A Option 001: 5 3/16 inches
- 85205B: 5 1/8 inches
- 85025D: 5 1/4 inches
- 85025E: 5 7/16 inches

Operating Environment

The detector will operate safely under the following conditions, but its performance is not necessarily warranted. See the specifications section for more information.

Temperature: 0° to +55 °C

Humidity: Up to 95%. Protect the detector from temperature extremes which can cause condensation.

Altitude: Up to 4,572 m (15,000 ft)

Accessories

The detectors come with a 2-meter cable. A 25-foot and 200-foot cable can be ordered separately. Table 6-1 lists the accessories that are available for use with these detectors.

Storage and Shipment

To keep your detector in proper working condition, keep the following suggestions in mind when storing or shipping it.

Environment

Store or ship the detectors in environments within the following limits:

Temperature: -25° to +75 °C

Humidity: Up to 95%. Protect the detector from temperature extremes which can cause condensation.

Altitude: Up to 4,572 m (15,000 ft)

Packaging

Use containers and materials identical or comparable to those used in factory packaging. If you ship the detector, follow these packaging instructions:

1. Wrap the detector in the original pouch and box. If they are not available, wrap the detector in heavy paper and use a strong shipping container.
2. Provide a firm cushion that prevents movement inside the container. Use a layer of shock-absorbing material around all sides of the detector.
3. Seal the shipping container securely.
4. Mark the shipping container *FRAGILE*.

Returning a Detector for Service

When you make an inquiry, either by mail or by telephone, refer to the detector by both model number and full serial number.

If you ship the detector to an Agilent office or service center, fill out a blue service tag (provided at the back of this manual), and include the following information:

1. Company name and address.
Do not use an address with a P.O. box number because products cannot be returned to a post office box.
2. The complete phone number of a technical contact person.
3. The complete model and serial number of the detector.
4. The type of service required (calibration, repair).
5. Any other information that could expedite service, such as failure condition or cause.

Installation

Refer to the following information when using the detector. Do not drop the detector or subject it to excessive mechanical shock.

Initial Inspection

1. Check the shipping container and packaging material for damage.
2. Check that the shipment is complete.
3. Check connector, cable, and detector body for mechanical damage.
4. Check the detector electrically:

Either perform the operator's check in Chapter 3, "Operation," or make a measurement in Chapter 4, "Performance Tests."

If any of the following conditions exist, notify your nearest Agilent office:

- Incomplete shipment.
- Mechanical damage or defect.
- Failed electrical test.

If you find damage or signs of stress to the shipping container or the cushioning material, keep them for the carrier's inspection. Agilent does not wait for a claim settlement before arranging for repair or replacement.


Electrostatic Discharge (ESD)

ESD can damage the highly sensitive circuits in this device; charges as low as 100 V can destroy a detector. ESD damage occurs most often as you connect or disconnect a device. Use this detector at a static-safe workstation and wear a grounding strap. *Never* touch the input connector center contacts or the cable contact pins.

Static-Safe Work Station

Figure 2-1 illustrates a static-safe station using two types of ESD protection that you can use either together or separately:

- A conductive table mat and wrist-strap combination.
- A conductive floor mat and heel-strap combination.



FORMAT48

Figure 2-1. Example of a Static-Safe Work Station

Static-Safe Practices

- Before cleaning, inspecting, or making a connection to a static-sensitive device or test port, ground yourself as far as possible from the test port.
- Discharge static electricity from a device before connecting it. Touch the device briefly (through a resistor of at least 1 M Ω) to either the outer shell of the test port, or another exposed ground. This discharges static electricity and protects test equipment circuitry.

Power Requirements

The scalar network analyzer supplies power for the detector.

Mating Connectors

Table 1-7 lists connector mechanical tolerances. *Microwave Connector Care* (part number 08510-90064) provides information on the proper maintenance, inspection, and gaging of connectors.

Connecting the Detector

1. The 85025A/B/D/E cables plug into the connectors on the front panel of the 8757 series scalar network analyzer. With the cable plug key downward, insert the multi-pin (DC) connector into the A input on the front panel of the analyzer.
2. To secure the DC connector in the analyzer, turn the *outer* shell clockwise.
3. Connect the RF input to the test device by turning the male connector's *outer* shell clockwise.

— |

| —


— |

| —

Operation

- Caution**
- Electrostatic discharge (ESD) can damage the highly sensitive circuits in this device; charges as low as 100 V can destroy your detector.
 - ESD damage occurs most often as you connect or disconnect a device. Use this detector at a static-safe workstation and wear a grounding strap. *Never* touch the input connector center contacts or the cable contact pins.
 - Do *not* apply more than +20 dBm RF power or more than ± 10 VDC to the detector. Higher power/voltage can electrically damage the detector.
 - Before you connect a RF cable to the detector, always discharge the static electricity that may have accumulated on the cable's outer conductor to instrument ground. This is most important if the cable is very long or connected to a large antenna.
 - Do not drop the detector or subject it to mechanical shock.
-

Features



1. **RF INPUT CONNECTOR.** This connector accepts the RF input signal. The RF input connector varies with the detector selected.
2. **DC CONNECTOR.** This connector supplies the necessary DC voltage for operation of the HP 85025A/B/D/E, and feeds the detector output signal to the network analyzer.

pc410a_e

Figure 3-1. Detector Features

Operating Theory

The 85025A/B/D/E can detect either unmodulated RF signals in DC mode or square wave amplitude modulated RF signals in AC mode. In either AC or DC detection mode, the detector provides a 27.778 MHz square wave signal for the analyzer to interpret and display.

In AC detection mode, an RF or microwave signal is amplitude modulated with a 27.778 MHz square wave. The detector demodulates (envelope detects) this signal to produce a 27.778 MHz signal with a peak-to-peak voltage that corresponds to the magnitude of the RF signal at the detector input.

In DC detection mode, no modulation is required. The detector diode in the 85025A/B/D/E converts the RF signal into an equivalent DC voltage. The detector chops the DC voltage at a 27.778 kHz rate, and this chopped signal is then amplified. The amplified signal simulates the signal produced by AC detection.

Measurement System Configuration

With an 8757 series scalar network analyzer using an 85025A/B/D/E detector, system configuration requires special attention. AC mode is the default state of the analyzer system and there are no further requirements to initiate a measurement. However, to enable DC mode operation, a series of keystrokes is required.

DC Detection mode

DC detection offers greater power measurement accuracy and the ability to characterize oscillators and modulation-sensitive devices. Figure 3-2 depicts a typical measurement setup for 0.01 to 50 GHz, using an Agilent 8350B sweep oscillator/RF plug-in as the source.

1. On the analyzer, press **PRESET**. Connect the detector(s).
2. DC detection mode must be selected. On the analyzer, press **SYSTEM** and select **MODE DC**.

When the **MODE DC** softkey is selected, the source's square wave modulation is automatically switched OFF.

Accurate DC Measurements

Zeroing the Detector

When you make DC detection measurements, it is important to perform this detector zeroing procedure to compensate for the effects of DC drift and temperature fluctuations. This zeroing procedure eliminates small DC voltages from the diode detector that would otherwise cause amplitude measurement errors at low (≤ 40 dBm) power levels. Zeroing also establishes the displayed noise level with no RF signal applied (the system's noise floor).

Autozero

Pressing the autozero softkey **AUTOZRO** switches OFF the source RF signal output and automatically zeros the detector.

The repeat autozero function softkey (REPT AZ ON/OFF) periodically repeats the autozero. You must use a GPIB interfaced sweeper to take advantage of this function because the analyzer must be able to switch OFF the RF output of the sweeper to perform the autozeroing.

Manual Zero

Manual zero, represented by the MANUAL softkey, is similar to zeroing a power meter.

1. Remove the RF signal from the detector's RF input.
2. On the analyzer, press MANUAL to perform the zeroing.

Refer to "Operation," in the *Agilent Technologies 8757C/E Scalar Network Analyzer Operating Manual* or *Agilent Technologies 8757D Scalar Network Analyzer Operating Manual* for detailed information on these and other softkeys.

In the DC mode, the 85025A/B/D/E is specified for absolute power level accuracy. In regard to these specifications, the following conditions apply:

- The equipment has had a 30 minute warmup period.
- The detector zeroing procedure has been performed.
- The offset has been adjusted with a calibrated 0 dBm, 50 MHz signal applied.
- Trace averaging is enabled on the analyzer at low power levels, as required.
- The source harmonics are below -40 dBc.
- The source SWR is 1.0.

To increase the accuracy of absolute power level measurements, select DET OFFSET to properly set the system response to a 0 dBm signal. After zeroing the detector, follow these steps to set the detector offset:

1. On the analyzer, press (CAL) DET OFFSET DET A (or the appropriate input port). Press (0) (dB). This ensures the 0 dB offset.
2. Connect the detector to the POWER REF output of a calibrated power meter, such as an Agilent 436A or 438A. Switch the POWER REF output ON.
3. Press (CURSOR) AUTOSCALE.
4. Press (CAL) MORE DET OFFSET.

5. Press **DET A** (or the appropriate input port) and use the entry keys to enter the value opposite in sign to the cursor reading being displayed. The display should now indicate a power reading of 0.00 dBm.

Note Pressing **PRESET** on the analyzer does not reset any DC OFFSET to zero, and the SAVE/RECALL registers do not save or recall the offset value(s).




Figure 3-2. Typical System Setup for 0.01 to 50 GHz Measurements

AC Detection Measurements

For the majority of measurements, AC detection is still the preferred method. AC detection offers greater sensitivity and immunity to noise and drift across time and temperature. AC detection amplitude measurements with this scalar network analyzer system require a modulation envelope. This envelope is provided through a 27.778 kHz square wave amplitude modulation of the RF test signal. Test set connections vary depending on the source. Figure 3-2 depicts a typical measurement setup. The 27.778 kHz modulation is supplied by the sweep oscillator.


Operator's Check

The following procedure provides a quick operational check of the 85025A/B/D/E detector.

Table 3-1. Equipment Required for Operator's Check

Description	Model/Part Number
Detector	85025A/B/D/E
GPIB cable	10833A/B/C/D
Bandpass filter	any within the frequency range of the sweeper
Sweep oscillator	any compatible with the 8757 series scalar network analyzer
BNC cables (3 required)	part number 8120-1839

Procedure



pc413a_e

Figure 3-3. Operator's Check Equipment Setup

1. Connect the equipment as shown in Figure 3-3, with the detector connected to input A on the analyzer. Form a thru connection by connecting the detector's RF input to the RF output of the source. Switch all of the instruments' line powers ON and allow them to warm up.
2. If the 8757 series analyzer's system interface is not engaged (the analyzer's status line displays `SYSINTF OFF`), press `(SYSTEM) MORE SWEEP MODE SYSINTF ON`.
3. On the analyzer, press `(PRESET)`. The analyzer's channel 1 should be active and measuring input A.
4. Adjust the STOP and START frequencies on the source to include the frequency range of the device under test (the bandpass filter).

5. Switch OFF the analyzer's channel 2 by pressing CHANNEL **2** twice.
6. On the analyzer, press **CURSOR** **MAX** to place the cursor at the maximum value of the trace. The CRSR value is displayed in the active entry area of the analyzer. Adjust the output power of the RF plug-in until the cursor value on the analyzer reads +16 dBm (this power level may not be attainable on all sources). This value is the upper limit of the dynamic range of the analyzer.
7. Press **DISPLAY** **MEAS**→**MEM** to store the trace in memory. The analyzer's message line displays CHAN 1 MEAS TO MEMORY.
8. Disconnect the detector from the RF OUTPUT of the source. Press **AVG** **AVG ON** to activate the averaging function. The averaging value will be 8. Wait a few seconds to allow the trace to settle.
9. Press **CURSOR** **MAX**. A cursor value of –55 dBm or lower should be displayed in the active entry area. This value represents the noise floor power level.
10. Insert the bandpass filter between the RF output of the source and the detector. Wait a few seconds to allow the trace to settle.
11. Press **CURSOR** **MAX** to find the trace maximum. The CRSR value displayed in the active entry area now represents the minimum insertion loss of your test device. Verify that the bandpass filter shape is as expected.
12. Press **SYSTEM** **MODE AC/DC** until the DC mode is activated. Allow the trace to settle. The trace should look similar to the trace observed in step 11, however the noise floor may be up to 5 dB higher. If the noise floor level has increased more than 5 dB, zero the detector.

If the Operator's Check Fails

Since the detector is measuring the output of an external source, problems may be due to the source, rather than the detector. Verify that the source output as well as the bandpass filter waveform are accurate. Note that bandpass filters can vary considerably from unit to unit.

If the average noise floor is not below –50 dBm in DC mode, try zeroing the detector. The instructions for zeroing the detector can be found earlier in this chapter.

Performance Tests

Use the procedures in this chapter to test the detector's electrical performance to the specifications listed in "General Information." None of these tests require access to the detector's interior.

To completely test each detector, three tests are required:

1. return loss
2. frequency response
3. absolute power/dynamic accuracy

Note For information on automating the Frequency Response Performance Test, see the chapter titled "Program Listing."

Equipment Required

Preceding each test is a list that describes the equipment required to perform that particular test. You may substitute any equipment that meets the indicated critical specifications. See Table 4-1 for an overall list of required equipment.

Note Before you proceed with the performance tests, gage the input connector on the detector and enter the results in the *Performance Test Record* at the end of "Performance Tests." For instructions on gaging the detector's input connector, see "Detector Maintenance."

Table 4-1. Recommended Equipment

Description	Recommended Agilent Model
Equipment Common to all 85025A/B/D/E Detectors	
Scalar network analyzer	8757C/D/E
Sweep oscillator	8350B
Power meter	436A, 437B, or 438A
Equipment required for 85025A Detectors	
RF plug-in	83592C
Power sensor	8481A
Directional bridge	85027C
Shielded open	p/n 85032-20002
Short	11511A
Adapter type-N(m) to type-N(m)	p/n 1250-1475
Attenuator	8491B Option 010
Equipment required for 85025A Option 001 Detectors	
RF plug-in	83592C
Power sensor	8481A Option 001
Directional bridge	85027A
Calibrated open/short	p/n 85021-60001
Adapter type-N(m) to type-N(m)	p/n 1250-1475
Attenuator	8492A Option 010
Equipment required for 85025B Detectors	
RF plug-in	83595C
Power sensor	8485B
Directional bridge	85027B
Calibrated open/short	p/n 85037-60001
Attenuator	8493C Option 010
Equipment required for 85025D Detectors	
Sweep oscillator	8350B
RF plug-in	83597B
Power sensor	8487A
Directional bridge	85027D
Open 2.4 mm(m)	85141A
Short 2.4 mm(m)	85140A
Adapter 2.4 mm(f) to 2.4 mm(f)	11900B
Attenuator	8490D Option 010

Table 4-1. Recommended Equipment (continued)

Description	Recommended Model
Equipment required for 85025E Detectors	
RF plug-in	83595C
Power sensor	8485A
Directional bridge	85027B
Calibrated open/short	p/n 85037-60001
Attenuator	8493C Option 010

Return Loss Performance Test

This performance test uses an 8757C/D/E scalar network analyzer system to measure the return loss of the detector.

85025A/B/D/E Return Loss Performance Test Procedure

Specifications

Specifications apply at a temperature range of 25 °C ±5 °C. For the detector's return loss specifications, refer to Tables 1-2 through 1-6.

Description

The return loss of the 85025A/B/D/E can be measured using the test system described in this procedure. The test setup is calibrated using an open/short to minimize frequency response and phasing errors. The detector under test is then connected to the TEST PORT of the bridge, and its return loss is measured on the analyzer.

The return loss should be greater than the limits listed in Tables 1-2 through 1-6. There is a certain amount of measurement uncertainty in any scalar network analyzer system. The return loss uncertainty for each detector is given in Tables 4-2 through 4-4. Conformance to specification cannot be assured unless the return loss of the detector is equal to the specified return loss plus the measurement uncertainty. Failure to meet specification cannot be proven unless measured return loss equals the specified return loss minus the measurement uncertainty. The measurement uncertainty is based on the worst case specifications for the test devices in the measurement.

Example A:

The specified return loss for an 85025A detector at 1 GHz is 20 dB. The measurement uncertainty is ± 1.4 dB.

If the detector's measured return loss is 20 dB + 1.4 dB (21.4 dB or higher), the detector is definitely within specification.

If the measured return loss is within the specified return loss plus-or-minus the measurement uncertainty, the detector may or may not be within specification. One way to reduce the measurement uncertainties is to measure the detector using a vector network analyzer.

Example B:

If the 85025A in Example A measures 20 dB ± 1.4 dB (i.e. from 18.6 to 21.4 dB), it cannot be determined if it is or is not within specification.

If the measured return loss is less than the specified return loss minus the measurement uncertainty, the detector is definitely out of specification.

Example C:

If the 85025A in Example A and B measured less than 20 dB $- 1.4$ dB (18.6 dB or less), the detector is definitely out of specification.

The three main sources of error in these measurements come from:

- Bridge directivity
- Source match of the bridge
- Dynamic accuracy of the analyzer

The first two vary with frequency while dynamic accuracy varies with the measured return loss amplitude.

Tables 4-2, 4-3, and 4-4 show measurement uncertainty for the 85025A/B, 85025D, and 85025E detectors, respectively. The tables assume you are using the corresponding connector-compatible Agilent directional bridges. For example, an 85027A bridge with an 85025A detector.

Table 4-2.
85025A/B Return Loss with Measurement Uncertainty

Model	Specified Return Loss and Measurement Uncertainty vs Frequency			
	0.01 to 0.04 GHz	0.04 to 4.0 GHz	4.0 to 18 GHz	18 to 26.5 GHz
85025A	10 ± 0.6 dB	20 ± 1.4 dB	17 ± 1.1 dB	
85025A Option 001	10 ± 0.6 dB	20 ± 1.1 dB	17 ± 0.7 dB	
85025B	10 ± 0.6 dB	20 ± 1.1 dB	17 ± 0.7 dB	12 ± 0.9 dB

Table 4-3.
85025D Return Loss with Measurement Uncertainty

Specified Return Loss and Measurement Uncertainty vs Frequency					
0.01 to 0.04 GHz	0.04 to 0.1 GHz	0.1 to 14 GHz	14 to 34 GHz	24 to 40 GHz	40 to 50 GHz
10 ± 0.5 dB	20 + 2, -1.5 dB	23 + 3, -2 dB	20 + 3, -2 dB	15 + 2, -1.5 dB	9 + 2, -1.5 dB

Table 4-4.
85025E Return Loss with Measurement Uncertainty

Specified Return Loss and Measurement Uncertainty vs Frequency			
0.01 to 0.04 GHz	0.04 to 0.1 GHz	0.1 to 25 GHz	25 to 26.5 GHz
10 ± 0.5 dB	20 ± 1 dB	25 ± 2 dB	23 ± 1.8 dB

Note An 85027A/B/C/D directional bridge (depending on the connector type) is used to measure the detector return loss.

Return Loss Measurement




Figure 4-1. Setup for 85025A/B/D/E Return Loss Test

1. Connect the equipment as shown in Figure 4-1, leaving the directional bridge TEST PORT unconnected.
2. Press **PRESET** on the analyzer. This will reset both the network analyzer and the sweep oscillator (the source).
Allow 30 minutes for warmup.
3. On the source, press **START** **10** **MHz** **STOP** **40** **MHz**.
4. On the RF plug-in, press **POWER LEVEL** and adjust the output power with the power level knob for a -3 dBm power level indication.

5. On the analyzer, press **CHAN 2 OFF** to switch channel 2 OFF. Press **FUNCTION** **(MEAS)** **A**. Press **FUNCTION** **(REF)** **REF POSN**, then use the step keys or the knob to move REF POSN one line down from the top of the CRT graticule.
6. On the analyzer, press **FUNCTION** **(SCALE)** **AUTO SCALE**.

Calibrating the Scalar Network Analyzer

7. On the analyzer, press **FUNCTION** **(CAL)** **SHORT/OPEN**. Follow the directions (prompts) appearing on the CRT:
 - a. Connect the short to the TEST PORT of the directional bridge, then press **STORE SHORT** on the analyzer. Remove the short.
 - b. Connect the open to the TEST PORT of the directional bridge, then press **STORE OPEN** on the analyzer. Remove the open.

The CRT will display **SHORT/OPEN CAL SAVED IN CH1 MEM**.

8. Press **FUNCTION** **(DISPLAY)** **MEAS-MEM**.
9. Connect the detector to be tested to the TEST PORT of the directional bridge. On the analyzer, press **FUNCTION** **(CURSOR)** and turn the analyzer's front panel knob to read the highest value (worst case return loss).

Record the worst case value in the space provided on the *Performance Test Record*, located at the end of this chapter.

Return Loss from 40 MHz to 18 GHz (to 26.5 GHz for 85025B/E only, to 50 GHz for 85025D only)

10. On the source, reset the START/STOP frequencies by pressing **(START)** **(40)** **(MHz)**. Next, for an 85025A detector, press **(STOP)** **(18)** **(GHz)**, for an 85025B/E detector, press **(26.5)** **(GHz)**, or for an 85025D detector, press **(50)** **(GHz)**.
11. On the analyzer, ensure **FUNCTION** **MEAS POWER A** is still active. Remove the detector.
12. It will be necessary to recalibrate the analyzer, since a new range of frequencies has been selected for measurement. Repeat the steps in the section titled "Calibrating the Measurement System."

Ensure that channel 1 display MEAS-MEM is active. **MEAS-MEM** is highlighted when active.

13. Connect the detector to the TEST PORT of the directional bridge.
14. On the analyzer, press FUNCTION **SCALE** **5** **dB**. The CRT display should be somewhat similar to Figure 4-2.




Figure 4-2. 85025A/B/E Return Loss 0.04 GHz to Maximum Frequency

15. On the analyzer, press **CURSOR**. Use the cursor to find the highest trace value in each specification range.

Note A specification range is a range of frequencies which have the same return loss specification. For example, the 85025A specification from 0.4 to 4 GHz.

16. Record each value in the *Performance Test Record*.

This completes the 85025A/B/D/E return loss performance test procedure.

If This Test Fails

Check the detector's input connector to make sure there is no damage. Open the detector's case and check the connection between the input connector and the PC board. Check that the detector is connected securely to the front panel of the analyzer. Replace the connector if necessary. If the detector still fails, refer to the "Service" chapter for more troubleshooting information.

Frequency Response Performance Test

Description

The frequency response of the 85025A/B/D/E detector is specified as the maximum peak-to-peak deviation from a constant input signal level of -10 dBm, as measured over the specified frequency range. At Agilent Technologies, frequency response is measured with the use of an automated test station traceable to the U.S. National Institute of Standards and Technology (NIST).

The "frequency response" specification for the 85025 family of detectors may more properly be called "frequency response flatness." The test for frequency response does not check absolute gain of the detector across its operating frequency range. It checks the relative variations in gain across the operating frequencies (for example, flatness).

To simplify the measurement procedure, frequency response is measured with a nominal -10 dBm signal applied. First, the source is characterized for frequency response using a calibrated power meter/sensor combination. Next, the DUT is characterized. Finally, a point-by-point difference is computed, plotted and compared to the specification window. Differences in the values recorded due to the different measurement scheme should be negligible.

The manual test described in this procedure has an approximate root sum of the squares (RSS) uncertainty ranging from ± 0.19 dB to ± 0.37 dB for 85025A/B/E detectors and ± 0.12 dB to ± 0.76 dB for 85025D detectors. This implies that a "good" detector, well within the limits of its specifications, could measure out of specifications.

This measurement is only an indication of the detector's response within these limits. If greater measurement accuracy is desired, a test system that minimizes

the sources of measurement uncertainty will be required. An error analysis of the sources of measurement uncertainty follows.

Table 4-5.
Approximate Error Analysis at 18 GHz for 85025A/B/E Detectors

Uncertainty	85025 A Option 001	85025 A	85025B	85025E
Power Sensor CAL Factor Uncertainties (RSS)	1.5%	1.5%	1.5%	1.5%
Mismatch between Attenuator and Power Sensor	5.2%	3.8%	2.1%	2.4%
Mismatch between Attenuator and Detector	5.9%	4.2%	3.3%	1.5%
Miscellaneous System Errors	1.1%	1.1%	1.1%	3.2%
RSS Calculation	8.1%	6.0%	4.3%	4.5%
Total RSS Uncertainties Expressed in dB	+0.34 to -0.37	+0.25 to -0.27	+0.18 to -0.19	+0.19 to -0.20

Table 4-6.
Approximate Error Analysis for the 85025D Detector

Uncertainty	26.5 GHz	40 GHz	50 GHz
Power sensor CAL factor approximate uncertainties	1.5%	2%	2.5%
Mismatch between attenuator and power Sensor	1.3%	3.8%	7.4
Mismatch between attenuator and detector	1.4	4.2	14.3
Miscellaneous system errors	1.1%	1.1%	1.1%
RSS calculation	2.7%	6.1%	16.3%
Total RSS uncertainties expressed in dB	+0.12 to -0.12	+0.26 to -0.27	+0.64 to -0.76

Uncertainties are smaller at lower frequencies. The error analysis is done assuming the power sensor, attenuator, and DUT all mate without the use of adapters. A standard 85025A is used with an 8481A and 8491B. An 85025A Option 001 detector is used with an 8481A Option 001 and 8492A. The 85025B/E is used with an 8485A and 8493C. The 85025D is used with an 8487A and 8490D.

Equipment Required

Equipment Common to 85025A/B/D/E

Sweep oscillator.....	8350B
Scalar network analyzer.....	8757C/D/E
Power meter.....	436A

Additional Test Equipment Required for 85025A

RF plug-in.....	83592C
Power sensor.....	8481A
10 dB attenuator.....	8491B Option 010

Additional Test Equipment Required for 85025A Option 001

RF plug-in.....	83592C
Power sensor.....	8481A
10 dB attenuator.....	8492A Option 010

Additional Test Equipment Required for 85025B

RF plug-in.....	83592C
Power sensor.....	8485A
10 dB attenuator.....	8493C Option 010
Adapter, type-N(m) to 3.5 mm(f).....	part number 1250-1744

Additional Test Equipment Required for 85025D Only

RF plug-in.....	83597B
Power Sensor.....	8487A
10 dB attenuator.....	8490D Option 010
Adapter, 2.4 mm(f) to 2.4 mm(f).....	11900B

Specifications

Specifications apply at a temperature range of 25 °C ±5 °C. For the detector's return loss specifications, refer to Tables 1-2 through 1-6.

85025A/B/D/E Frequency Response Performance Test Procedure

Note

For 85025D detectors:

While the detector is specified down to 10 MHz, the power sensor used in this procedure is only calibrated down to 50 MHz. If data is required below 50 MHz, characterize the power sensor to 10 MHz, or use an additional power sensor which covers this frequency range, and correlate the results with the data above 50 MHz.

Configuring the System

1. Connect the equipment as shown in Figure 4-3, with nothing connected to the attenuated output of the source. Switch ON all equipment and allow 30 minutes for warmup.

2. On the power meter, press **[dBm]** mode.

Zero and calibrate the power meter. If you are unsure of how to do this, refer to the power meter's Operating and Service Manual.

[RANGE HOLD] and **[POWER REF]** should remain out.

Set the CAL FACTOR % dial on the power meter to the value indicated for 50 MHz on the power sensor CAL FACTOR chart.

3. On the analyzer, press **[PRESET]** **[CHANNEL]** **CHAN 2 OFF** **INSTRUMENT STATE** **[SYSTEM]** **MODE DC**.

4. On the analyzer, zero the detector by pressing **DC DET ZERO** **MANUAL** **CONT**.

When the zero is complete, the display will indicate: **MANUAL ZERO COMPLETE**.

5. Connect the power meter/sensor to the attenuated RF output.

6. On the source/RF plug-in, press **[CW]** **[5]** **[0]** **[MHz]**.

7. Adjust the power level for an indication of -10 dBm on the power meter.
Do *not* readjust the power level for the remainder of this test.

Characterizing the Source

8. On the source, press **CW** and enter the desired test frequency as indicated on the *Performance Test Record* located at the end of this chapter. For example: **CW** **0.01** **GHz**.
9. Using the CAL FACTOR chart on the power sensor, set the CAL FACTOR % dial on the power meter to the value indicated for the test frequency as needed. Use the nearest frequency value.
10. Note the reading on the power meter.
11. Record this value and the test frequency in the *Performance Test Record*.
12. Repeat steps 9 and 10 until the source is characterized to your satisfaction.

Characterizing the Detector

13. Disconnect the power meter/sensor.
14. On the analyzer, zero the detector by pressing **DC DET ZERO** **AUTOZRO**.
When the zero is complete, the display will indicate: **AUTOZERO COMPLETED**.
15. Connect the detector between the attenuated output of the source and INPUT A of the analyzer.
16. On the analyzer, press **CHAN 2 OFF** to switch channel 2 OFF. Press **FUNCTION** **CURSOR** to switch the cursor ON.
17. On the source, press **CW** and enter the value of the first test frequency. Remember to use only the test frequencies used in steps 9 through 11.
Note the value indicated by the analyzer's cursor display and record it in the *Performance Test Record*.
Repeat this step until all of the same frequency points have been characterized.

Computing the Maximum Error

18. Using the values recorded in steps 11 and 17, subtract the value in step 11 from the value in step 17 for each of the test frequencies. Record the difference in the space provided on the *Performance Test Record*.

Now use these values to plot a point-to-point variation curve on the graph provided in the *Performance Test Record*.

19. Read the “frequency response” limits from the appropriate specifications table in Chapter 1. Plot these limits onto the graph with the point-to-point variation curve.
20. The detector is considered to “pass ” this test if the variation curve is contained within the limits, OR if an offset can be applied to the variation curve so that it is contained within the limits. (In other words, can the variation curve be shifted up or down to make it fit within the limits?)

This completes the procedure for measuring frequency response.

Power Accuracy Performance Test

Specifications

Refer to Table 1-2 through Table 1-6. Specifications apply at 25 °C ±5 °C.

Description

Power accuracy is measured with consideration given to the following conditions (the order is not important):

- Power accuracy is measured at 50 MHz.
- If in DC mode, autozero has been performed (not necessary for AC mode).
- All equipment has been allowed to warmup for 30 minutes.
- Source harmonics are below –40 dBc.
- This performance test includes mismatch effects. Specifications assume no mismatch.
- Trace averaging should be used on the analyzer as required.
- Offset active and adjusted with a calibrated 0 dBm, 50 MHz signal applied (DC mode only).

The recommended method for testing power accuracy is to use an Agilent 8116A pulse/function generator as the source. An 8116A provides the amplitude

necessary to check the detector to its full specifications. Both AC and DC modes must be tested to verify the performance specifications of the detector.

Note that the DC mode test is an “absolute” measurement, requiring the use of a calibrated power meter to set a level of 0 dBm. Using the 436A as a calibrated 0 dBm, 50 MHz source introduces a maximum measurement uncertainty of ± 0.07 dB.

AC power accuracy testing is done with the 8116A modulated by an 11665B.

An alternate procedure is provided using an 8350B with an RF plug-in as the source. This method does not test the detector to its full specifications (tests to +10 dBm) and should not be used when traceability to the National Institute of Standards and Technology (NIST) is required.

Equipment Required

Pulse/function generator	8116A
Scalar network analyzer	8757C/D/E
50 MHz bandpass filter	part number 08757-80027
3 dB attenuator	8491B
Calibrated 10 dB step attenuator	355D Option 001/H88
Calibrated 1 dB step attenuator	355C Option 001/H88
Modulator (AC mode only)	11665B
Power meter (DC mode only)	436A/438A
Adapters	as required

Note Calibrated attenuation is used in the power accuracy calculations below. Calibrated step attenuators include a calibration report at 50 MHz to improve measurement accuracy. The report lists the actual attenuation of each step at one frequency of interest. The calibration report may be ordered as an option with the step attenuators when purchased or performed as a service afterwards.

Procedure

Absolute Power Accuracy in DC Mode Performance Test

1. Connect the equipment as shown in Figure 4-4. Do not connect detector to attenuator output. Do not use the modulator for this test (AC mode only). Switch all the equipment ON and allow 30 minutes warmup time.




Figure 4-4. Absolute Power Accuracy Test Setup

2. For each of the power levels specified in columns 2 and 3 of the *Performance Test Record*, record the required calibration data of the step attenuators (10 dB step attenuator data in column 4 and 1 dB step attenuator data in column 5).
3. Calculate the calibrated power level for each power level and record this value in column 6 CAL PWR LVL. An example follows:

1. Nominal PWR LVL	2. Nominal 10 dB STEP ATTEN	3. Nominal 1 dB STEP ATTEN	4. CAL ATTEN	5. CAL ATTEN	6. CAL PWR LVL
0	10	6	10.02	6.01	-0.03

$16 \text{ dBm} - (\text{column 4} + \text{column 5}) = \text{CAL PWR LVL}$

$16 \text{ dBm} - (10.02 \text{ dB} + 6.01 \text{ dB}) = -0.03 \text{ dBm}$

4. On the 8116A, set a frequency of 50 MHz by pressing Sine Function (P) Duty (DTY). Using the VERNIER rocker keys adjust for a 50% duty cycle display. Press Frequency (FRQ). Using the VERNIER and RANGE rocker keys adjust for a 50 MHz display. Select normal operation. LED NORMAL ON.
5. On the analyzer, press (PRESET) CHAN 2 OFF. Switch continuous wave ON and select DC mode for the detector by pressing INSTRUMENT STATE (SYSTEM) MORE SWEEP MODE CW ON, INSTRUMENT STATE (SYSTEM) MODE DC.
6. Configure the analyzer inputs and perform a manual DC zero. On the analyzer, press FUNCTION (CAL) CONFIG SYSTEM DC DET ZERO MANUAL CONT.
7. Perform the detector offset calibration. On the analyzer, press DET OFFSET A (0) (dB). This ensures 0 dB of offset.
8. Connect the detector to the power meter POWER REF output.
9. Switch POWER REF output ON.

10. On the analyzer, perform the following:
 - a. Press FUNCTION **(SCALE)** **AUTO SCALE** FUNCTION **(CURSOR)**. Note the reading for use in step b.
 - b. Press FUNCTION **(CAL)** **DET OFFSET A**. Using the ENTRY keys, enter the value opposite in sign to the reading noted above.
 Example: CRSR = +.45 dBm
 Press **(-)** **(.)** **(4)** **(5)** **(dB)**.
 - c. Press FUNCTION **(CURSOR)**. The display should indicate a power level of 0.00 dBm. If not, repeat the detector DC ZERO and OFFSET CALIBRATION (steps 6 and 7) until a 0.00 dBm power level is obtained.
11. On the step attenuators, set the attenuators for a total of 16 dB attenuation. Set the 355C to 6 dB and set the 355D to 10 dB.
12. Connect the DUT to the attenuated output.
13. On the 8116A enable the output by pressing **(DISABLE)**. The DISABLE LED should deactivate.

 Refer to the *Performance Test Record* at the end of this chapter for the calibrated power level computed at nominal 0 dBm. Use the VERNIER rocker keys to adjust the output power to the CAL PWR LVL.
14. Note the cursor value displayed on the CRT. Record this value in the space provided in column 7 of the *Performance Test Record*.
15. Set both attenuators to 0 dB attenuation. Note and record the cursor value.
16. Set the attenuators for the next Nominal PWR LVL (dBm).

Note For nominal power levels of –16 dBm and below, use a combination of AVERAGING ON and SMOOTHING ON to reduce trace noise and obtain a stable reading. Refer to the *Performance Test Record* for the specified AVERAGING FACTOR. Allow for settling time after resetting the attenuator(s).

17. When the cursor reading has stabilized, note and record the value in column 7.
18. Repeat steps 13 and 14 for each nominal power level listed on the *Performance Test Record*.

19. Calculate the Dynamic Accuracy Error as follows:

$$\text{Dynamic Accuracy Error} = \text{MEAS PWR LVL} - \text{CAL PWR LVL}$$

Include and preserve signs in this calculation. Enter this value in the Dynamic ACCY Error (dBm), column 8, of the *Performance Test Record*.


Dynamic Accuracy in AC Mode Performance Test

20. Connect equipment as shown in Figure 4-4. Connect the modulator's DRIVE INPUT to the analyzer's rear panel MODULATOR OUTPUT. Connect the DUT to the attenuated output. Switch the equipment ON and allow 30 minutes for warmup.
21. Record and calculate the data as necessary into columns 4, 5, and 6 of the *Performance Test Record*. If the attenuators used in this test are the same as the ones used in steps 2 and 3, copy the data from the Absolute Power Accuracy in DC Mode *Performance Test Record* (columns 4, 5, and 6).
22. On the analyzer, press **[PRESET]** **CHAN 2 OFF** **FUNCTION** **[CURSOR]**.
23. Set the 10 dB step attenuator to 10 dB and set the 1 dB step attenuator to 6 dB.
24. On the 8116A, set a frequency of 50 MHz by pressing Sine Function **[P]** Duty **[DTY]**. Using the VERNIER rocker keys adjust for a 50% duty cycle display. Press Frequency **[FRQ]**. Using the VERNIER and RANGE rocker keys adjust for a 50 MHz display.
Select normal operation. LED NORMAL on.
Refer to the *Performance Test Record* for the CAL PWR LVL computed at nominal 0 dBm. Use the VERNIER rocker keys to adjust the output power to the CAL PWR LVL as displayed by the CURSOR on the analyzer.
25. Note and record on the *Performance Test Record* the cursor value displayed.
26. Set both attenuators to 0 dB. Note and record the cursor value.
27. Set the attenuators for the next Nominal PWR LVL. Continue the procedure as outlined in steps 24 through 26 of the DC Mode Test for each of the Nominal PWR LVLs listed on the *Performance Test Record*.
This completes the procedure for measuring dynamic accuracy.

Power Accuracy, Alternate Procedure Using an 8350B (+ 10 dBm maximum)

Alternate Equipment

Sweep oscillator	8350B
RF plug-in	83592B
Scalar network analyzer	8757C/D/E
50 MHz bandpass filter	Part Number 08757-80027
3 dB attenuator	8491B
Calibrated 10 dB step attenuator	355D Option 001/H88
Calibrated 1 dB step attenuator	355C Option 001/H88
Power meter (DC mode only)	436A or 438A
Adapters	as required



pc420a_e

Figure 4-5. Power Accuracy Alternate Test Setup

Procedure

1. Connect the equipment as shown in Figure 4-5. Allow 30 minutes warmup time.

2. For each of the nominal power levels specified in the *Performance Test Record* (Alternate Procedure), calculate the Calibrated Power Level. Refer to steps 2 and 3 in the Absolute Power Accuracy in DC Mode Performance Test for details. Change all +16 dBm references to +10 dBm.

Absolute Power Accuracy in DC Mode, Alternate Procedure

3. Do not connect the detector to the attenuated output.
4. On the analyzer, press (PRESET) CHAN 2 OFF INSTRUMENT STATE (SYSTEM) (MODE DC). Next, press (CAL) CONFIG SYSTEM DC DET ZERO MANUAL CONT DET OFFSET A (0) (dB).
5. Connect the detector to the power meter POWER REF output.
6. Switch the POWER REF output ON.
7. On the analyzer, press (SCALE) AUTO SCALE (CURSOR). Note the reading.

On the analyzer, press (CAL) DET OFFSET A. Using the ENTRY keys, enter the value opposite in sign to the reading noted above.
8. On the analyzer, press (CURSOR). The display should indicate a power level of 0.00 dBm. If not, repeat the detector zero and offset calibration until a 0.00 dBm power level is obtained (steps 4 through 7).
9. Set the 10 dB step attenuator to 10 dB.
10. Connect the DUT to the attenuated output.
11. On the source, switch square wave modulation OFF. Set a CW frequency of 50 MHz.
12. Refer to the *Performance Test Record* for the CAL PWR LVL computed at nominal 0 dBm. Adjust the output power to the CAL PWR LVL.
13. Note the cursor value displayed on the CRT. Record this value in column 7 of the *Performance Test Record*.
14. Set the attenuators for the next nominal PWR LVL. Continue the procedure as outlined in steps 9 through 13 of the Absolute Power Accuracy in DC Mode Performance Test for each of Nominal PWR LVLs listed on the *Performance Test Record* (Nominal -55 dBm is for AC Test only).

Dynamic Accuracy in AC Mode, Alternate Procedure

15. Repeat steps 1 and 2 of Absolute Power Accuracy in DC Mode Performance Test, Alternate Procedure. Connect the DUT to the attenuated output.
16. Preset the analyzer. Switch OFF channel 2. Switch ON the cursor.
17. Set the 10 dB step attenuator to 10 dB.
18. On the source set CW to 50 MHz.
19. Refer to the *Performance Test Record* for the CAL PWR LVL computed at nominal 0 dBm. Adjust the output power to the CAL PWR LVL.
20. Note and record the displayed cursor value.
21. Set both attenuators to 0 dB. Note and record the cursor value.
22. Set the attenuators for the next nominal PWR LVL. Continue the procedure as outlined in steps 17 through 19 of the Absolute Power Accuracy in DC Mode Performance Test for each of the nominal PWR LVLs listed on the *Performance Test Record*.

This completes the power accuracy alternate procedure.

Performance Test Record

85025 A/B/D/E Detector Performance Test Record (1 of 9)

Test Facility _____	Report Number _____
_____	Date _____
_____	Customer _____
_____	Tested by _____
Model _____	Ambient temperature _____ °C
Serial Number _____	Relative humidity _____ %
Options _____	
Special Notes	

85025A/B/D/E Detector Performance Test Record (2 of 9)

Model _____	Report Number _____	Date _____	
Test Equipment Used	Model Number	Trace Number	Cal Due Date
1. _____	_____	_____	_____
2. _____	_____	_____	_____
3. _____	_____	_____	_____
4. _____	_____	_____	_____
5. _____	_____	_____	_____
6. _____	_____	_____	_____
7. _____	_____	_____	_____
8. _____	_____	_____	_____

85025A/B/D/E Detector Performance Test Record (3 of 9)

Test Description	Specification	Measured Results	Measurement Uncertainty ¹
85025A/B Return Loss			
10 MHz to 40 MHz	≥ 10 dB	_____	±1.4 dB
0.04 GHz to 4.0 GHz	≥ 20 dB	_____	±1.1 dB
4 GHz to 18 GHz	≥ 17 dB	_____	±1.4 dB
18 GHz to 26.5 GHz ²	≥ 12 dB	_____	±1.1 dB
85025D Return Loss			
10 MHz to 40 MHz	≥ 10 dB	_____	±0.5 dB
0.04 GHz to 0.10 GHz	≥ 20 dB	_____	+2, -5 dB
0.10 GHz to 14 GHz	≥ 23 dB	_____	+3, -2 dB
14 GHz to 34 GHz	≥ 20 dB	_____	+3, -2 dB
34 GHz to 40 GHz	≥ 15 dB	_____	+2, -1.5 dB
40 GHz to 50 GHz	≥ 9 dB	_____	+2, -1.5 dB
85025E Return Loss			
10 MHz to 40 MHz	≥ 10 dB	_____	±1.4 dB
0.04 GHz to 0.1 GHz	≥ 20 dB	_____	±1.1 dB
0.1 GHz to 25 GHz	≥ 25 dB	_____	±1.4 dB
25 GHz to 26.5 GHz	≥ 23 dB	_____	±1.1 dB



1 Using the equipment and procedures documented in this manual.

2 85025B only

85025A/B/D/E Detector Performance Test Record (4 of 9)

Frequency Response				
Test Frequency		Source Step 11	Detector Step 17	Difference Step 18
Recommended	Actual			
0.01 GHz				
2.00 GHz				
4.00 GHz				
6.00 GHz				
8.00 GHz				
10.0 GHz				
12.0 GHz				
14.0 GHz				
16.0 GHz				
18.0 GHz				
20.0 GHz				
21.0 GHz				
22.0 GHz				
23.0 GHz				
24.0 GHz				
25.0 GHz				
26.0 GHz				
26.5 GHz				
28.0 GHz				
30.0 GHz				
32.0 GHz				
34.0 GHz				
36.0 GHz				
38.0 GHz				
40.0 GHz				
42.0 GHz				
44.0 GHz				
46.0 GHz				
48.0 GHz				
50.0 GHz				

85025A/B/D/E Detector Performance Test Record (5 of 9)



85025A/B/D/E Detector Performance Test Record (6 of 9)

Absolute Power Accuracy in DC Mode								
1. Nominal PWR LVL (dBm)	2. Nominal 120 dB ATTEN Setting (dB)	3. Nominal 12 dB ATTEN Setting (dB)	4. CAL ATTEN (120 dB ATTN)	5. CAL ATTEN (12 dB ATTN)	6. CAL PWR LVL (dBm) (16 dBm - CAL ATTN)	7. MEAS PWR LVL (Cursor) (dBm)	8. Dynamic ACCY Error (dBm)	9. Upper Limit ¹ (85025A/B/D/E unless noted otherwise)
0	10	6					REF	REF
16	0	0						0.70 ² 1.0 ³
13	0	3						0.61 ² 0.76 ³
10	0	6						0.52
6	10	0						0.40
3	10	3						0.40
0	10	6					REF	REF
-3	10	9						0.40
-6	20	2						0.40
-10	20	6						0.40
-13	20	9						0.40
-16 ⁴	30	2						0.40
-20 ⁴	30	6						0.40
-25 ^{4, 5}	40	1						0.40
-30 ^{4, 5}	40	6						0.40
-35 ^{4, 6}	50	1						0.40
-40 ^{4, 6}	50	6						0.70
-45 ^{4, 7}	60	1						1.0
-50 ^{4, 7}	60	6						1.3

1 Upper limit does not include measurement uncertainties.

2 85025A/B only

3 85025D/E only

4 Smoothing ON

5 Averaging ON, Averaging Factor=4

6 Averaging ON, Averaging Factor=8

7 Averaging ON, Averaging Factor=8

85025A/B/D/E Detector Performance Test Record (7 of 9)

Dynamic Accuracy in AC Mode								
1. Nominal PWR LVL (dBm)	2. Nominal 120 dB ATTEN Setting (dB)	3. Nominal 12 dB ATTEN Setting (dB)	4. CAL ATTEN (120 dB ATTN)	5. CAL ATTEN (12 dB ATTN)	6. CAL PWR LVL (dBm) (16 dBm - CAL ATTN)	7. MEAS PWR LVL (Cursor) (dBm)	8. Dynamic ACCY Error (dBm)	9. Upper Limit ¹ (85025 A/B/D/E unless noted otherwise)
0	10	6					REF	REF
16	0	0						0.70 ² , 1.0 ³
13	0	3						0.61 ² , 0.76 ³
10	0	6						0.52
6	10	0						0.40
3	10	3						0.40
0	10	6					REF	REF
-3	10	9						0.40
-6	20	2						0.40
-10	20	6						0.40
-13	20	9						0.40
-16 ⁴	30	2						0.40
-20 ⁴	30	6						0.40
-25 ^{4, 5}	40	1						0.40
-30 ^{4, 5}	40	6						0.40
-35 ^{4, 6}	50	1						0.40
-40 ^{4, 6}	50	6						0.70
-45 ^{4, 7}	60	1						1.0
-50 ^{4, 7}	60	6						1.3
-55 ^{4, 8}	70	1						1.6

1 Upper limit does not include measurement uncertainties.

2 85025A/B only

3 85025D/E only

4 Smoothing ON

5 Averaging ON, Averaging Factor=4

6 Averaging ON, Averaging Factor=8

7 Averaging ON, Averaging Factor=8

8 Averaging ON, Averaging Factor=32

85025A/B/D/E Detector Performance Test Record (8 of 9)

Alternate Tests								
1. Nominal PWR LVL (dBm)	2. Nominal 120 dB ATTEN Setting (dB)	3. Nominal 12 dB ATTEN Setting (dB)	4. CAL ATTEN (120 dB ATTN)	5. CAL ATTEN (12 dB ATTN)	6. CAL PWR LVL (dBm) (16 dBm - CAL ATTN)	7. MEAS PWR LVL (Cursor) (dBm)	8. Dynamic ACCY Error (dBm)	9. Upper Limit ¹
0	10	0					REF	REF
10	0	0						0.52
6	0	4						0.40
3	0	7						0.40
0	10	0					REF	REF
-3	10	3						0.40
-6	10	6						0.40
-10	20	0						0.40
-13	20	3						0.40
-16	20	6						0.40
-20	30	0						0.40
-25 ²	30	5						0.40
-30 ²	40	0						0.40
-35 ^{2,3}	40	5						0.40
-40 ^{2,3}	50	0						0.70
-45 ^{2,4}	50	5						1.0
-50 ^{2,4}	60	0						1.3
-55 ^{2,5}	60	5						1.6
-50 ^{2,5}	60	6						1.3
-55 ^{2,6}	70	1						1.6

1 Upper limit does not include measurement uncertainties.

2 Smoothing ON

3 Averaging ON, Averaging Factor= 4

4 Averaging ON, Averaging Factor= 8

5 Averaging ON, Averaging Factor= 8

6 Averaging ON, Averaging Factor= 32

85025A/B/D/E Detector Performance Test Record (9 of 9)

RF Input Connector Mechanical Tolerances			
Connector	Minimum Recession (inches) ¹	Maximum Recession (inches)	Measured Recession (inches)
85025A Type-N male	0.207	0.210 ²	_____
85025A Option 001 Precision 7 mm ³	0.000	0.003	_____
Precision 7 mm Collet Resilience	Collet must spring back out after being depressed.		Pass _____ Fail _____
85025B/E Precision 3.5 mm male	0.000	0.003	_____
85025D Precision 2.4 mm male	0.000	0.002	_____

- 1 Minimum recession must NEVER be less than the minimum recession tolerance. If a connector fails this specification, immediately replace it. Such connectors will damage any connector mated to it.
- 2 The type-N gage calibration block zeros the gage at a 0.207 inch offset. Therefore, the 0.207 inch to 0.210 inch recession is displayed as 0.000 to 0.003 inches on the gage.
- 3 With collet removed.

— |

| —

— |

| —

Adjustments

There are no adjustments which can be performed on the 85025A/B/D/E.

— |

| —

— |

| —

Replaceable Parts

This chapter provides information on ordering replaceable parts. To order a part listed in Table 6-1:


- Quote the Agilent part number.
- Indicate the quantity required.
- Address the order to your nearest Agilent office.

To request information on a part that is not listed in Table 6-1, include the instrument model number, a description of the part, and its function. Address the inquiry to the nearest Agilent office.

How To Order Parts Fast


When you know which parts you need to repair the detector, contact Agilent Technologies. The parts specialists have direct online access to the replacement parts inventory corresponding to Table 6-1 in this guide.

Table 6-1. 85025A/B/D/E Replaceable Parts and Accessories



Item Number	Description	Agilent Part Number
1	Replacement 85025A Detector, Type-N	85025-69018
1	Replacement 85025A Option 001 Detector, 7 mm	85025-69019
1	Replacement 85025B Detector, 3.5 mm	85025-69020
1	Replacement 85025D Detector, 2.4 mm	85025-69024
1	Replacement 85025E Detector, 3.5 mm	85025-69025
2	85025A Connector Repair Kit, Type-N	85025-60021
2	85025A Option 001 Connector Repair Kit, 7 mm	85025-60022
2	85025B Connector Repair Kit, 3.5 mm	85025-60023
2	85025D Connector Repair Kit, 2.4 mm	85025-60026
2	85025E Connector Repair Kit, 3.5 mm	85025-60027
3	Label, Serial Number	Not Available
4	Label, Warning Max Input	85025-80001
5	Cable Assembly (W1)	85025-60003
	Cover, Plastic Half Body	85025-40001
	Screw-Machine M2.5 × 0.45; 4 mm-LG	0515-0061
	Washer-Lock 2.5 mm	2190-0583
	Cable Marker Kit	5061-1044
	Cable, 25 foot	11679A
	Cable, 200 foot	11679B

The module exchange program described here is a fast, efficient, economical method of keeping your Hewlett-Packard instrument in service.



*HP pays postage on boxes mailed in U.S.A.

pc426a-e

Figure 6-1. Module Exchange Program

— |

| —

— |

| —

Service

Caution This product is susceptible to damage from electrostatic discharge (ESD). When you perform any of the following procedures, wear a grounded static-strap and work at a static-safe work station.

If the detector fails electrically, order a replacement detector. Do not order the detector using its model number (85025A/B/D/E). Instead, use the replacement part number given in Table 6-1 that is referenced to the model number.

These detectors have the following replaceable items:

- The input connector.
- The cable assembly.


Repair

Removing the Covers

A small flat-blade screwdriver with a blade width no greater than 3.5 mm (1/8 inch) is required to perform this procedure.

Procedure

1. Place the detector so its narrow side is on a flat surface. Position it so that the RF connector is facing away from you. Refer to Figure 7-1.



pc49ae_c

Figure 7-1. Removing the Detector Covers

2. Hold the sides of the detector near the cable-end. Insert the screwdriver at a 45 degree angle between the side label and the raised edge of the plastic cover as shown in Figure 7-1. Make sure the screwdriver is inserted as far forward as possible on the detector.
3. Rotate the screwdriver about 90 degrees as shown, until the cover snaps apart.
4. Repeat steps 2 and 3, inserting the screwdriver approximately 2/3 of the way toward the cable-end of the detector, at the point shown in Figure 7-1.
5. Separate the plastic shell halves. If the cover does not separate easily, repeat steps 2, 3, and 4 on the other side of the detector.
6. Attach the covers to the replacement detector by snapping the halves together.

Replacing the Detector

1. Remove the plastic covers from the existing detector using the cover removal procedure.
2. Install the covers on the replacement detector.
3. Perform the operator's check described in "Operation."
4. If using a restored exchange replacement detector, return the defective detector using the packing material supplied. This should be done *immediately* in order to assure proper credit for the exchange.

Note If the existing plastic half cover is damaged or broken, replacement covers are available. The part number is provided in the replaceable parts list. Two are required for each detector.


Replacing the Cable Assembly

1. Order the cable assembly (Item 5 in Table 6-1).
2. Remove the outer covers using the cover removal procedure indicated earlier in this section.
3. Remove the two pozi-drive screws located at the cable end of the metal housing.

4. Slide the metal housing away from the RF connector to expose the printed circuit (P.C.) assembly completely.
5. Carefully de-solder all cable wires from the P.C. assembly.
6. With the P.C. assembly facing up, secure the detector frame in a vice. Be careful to set the detector so the P.C. assembly is NOT gripped.

Caution Over-tightening the vice causes the frame to bend.

7. Use a 7/16-inch open-ended wrench to unscrew the cable hex nut.
8. Remove the old cable assembly and remove the metal housing from the cable.
9. Carefully slide the metal housing onto the new cable assembly. Be sure to place the metal housing so the adjustment potentiometers are accessible.
10. Screw the cable assembly onto the detector frame.
11. Solder the wires of the new cable to the P.C. assembly (refer to Figure 7-2 for proper placement).
12. Ensure that all cable wires are securely connected to the assembly.
13. Slide the metal housing over the P.C. assembly and secure to the frame using pozi-drive screws.
14. Snap the outer covers back on.



pc425a_e

Figure 7-2. Cable Connections

Replacing the Connectors

Order the appropriate Connector Repair Kit (see Table 6-1) for your detector. Follow the instructions provided in the kit to replace the connectors.

— |

| —

— |

| —

Detector Maintenance

This chapter describes how to maintain your detector in proper working order. This includes:

- How to inspect the detector
- How to clean the input connector
- How to gage the input connector

Mechanical Inspection

This section provides a brief introduction to the fundamentals of proper connector care, which is as important to making good measurements as proper instrument calibration and adjustment.

This is not intended to be a comprehensive discussion of this vital topic. For information on connector care, refer to the *Microwave Connector Care Quick Reference Card*, part number 08510-90360.

Inspecting the Connectors

Visual and mechanical inspection of the connectors should be done periodically. If a bad connector is accidentally attached to a good connector, the good connector can be damaged. The time and expense involved in replacing detectors or other devices due to damaged connectors warrants caution.

A connector is bad if one of the following conditions exist:

- It fails the visual examination.
- It fails the mechanical examination.
- When attaching two connectors together, they do not mate smoothly.

Visual Examination

A careful visual inspection should be performed often on all system connectors. Vigilance can save money and ensure accurate measurements with your equipment.

Examine the connectors for such obvious problems as deformed threads, contamination, or corrosion, concentrating especially on the contacting surfaces. Look for burrs, scratches, rounded shoulders and similar signs of wear or damage. Any problem you can see is sufficient to cause degraded performance, and the detector must be cleaned or replaced.

Cleaning the Connectors

In harsh environments, the connectors might become dirty. To safely clean the connector, carefully brush or wipe dirt from the surface with a foam swab. It is safe to use trichlorotrifluoroethane (liquid Freon) sparingly as a cleaning solvent. However, it is not safe to use abrasives of any kind (such as pencil eraser) or any other solvent, because of damage to the thin metal plating or to the plastic dielectric supporting element.

Connector Cleaning Kit

When cleaning an RF connector, Agilent recommends that you use the Connector Cleaning Kit, part number 92193Z.

Gaging Connectors

Gaging connectors is necessary to ensure that they conform to mechanical tolerances. Out-of-tolerance connectors have poor electrical performance and could damage another connector mated to them. Refer to the *Microwave Connector Care Quick Reference Card*, part number 08510-90360.

Refer to Table 1-7, "Supplemental Characteristics," for 85025A/B/D/E connection tolerances.

Gaging Connectors to be Mated with the 85025A/B/D/E

It is important to gage connectors which will be used with the 85025A/B/D/E. The specifications for attaching connectors may vary, depending on the connector used and the application. However, if the following guidelines are used, you can avoid connector damage due to tolerance problems.

Type-N female

Gage any female device to be mated with the 85025A. When mated, a type-N connector pair must have separation between the tip of the female contact fingers and the shoulder of the male contact pin. Do not use female type-N connectors that have an inner rubber washer, they damage the male center conductor.

Precision 7 mm

Gage with the center conductor collet removed, refer to the tolerances given in the device's manual. There should be no protrusion of the center conductor in front of the outer conductor mating plane. Check the collet for proper spring action. Collets should only be removed with a collet extraction tool. Refer to the *Microwave Connector Care Quick Reference Card* for information.

Precision 3.5 mm female

Gage any female connector to be used with the 85025B/E. There should be no protrusion of the tip of the female contact fingers in front of the outer conductor mating plane.

— |

| —

— |

| —

Automated Program Listing

The information in this section is provided as a convenience only. It is intended to provide a starting point for making automated measurements and does not contain full error detection or extra enhancements. This program is not warranted to be uninterrupted or error free.

Automating the Frequency Response Test

The program listing at the end of this chapter allows you to automate the frequency response test of the detector. This test determines the frequency response at up to 401 points in just a few minutes. The program first measures the frequency response of the source (taking into account the frequency response of the power sensor), then measures the frequency response of the detector, and finally plots the difference between the two on the computer's CRT. This is identical to the manual version of the test and the results should be identical.

As written, this program will test detectors to 26.5 GHz. If you need results to 18 or 50 GHz, you must modify the START/STOP frequencies of the program.

The following equipment (.01 to 50 GHz) is recommended to run the program:

- HP 9000 series 200 or 300 computer with BASIC 2.0 or above
- A signal source: 83752A/B, 8360 Series, or 8350B with plug-in (depending on the operating frequency of the detector you are testing)

Note The 8360 Series and 83752A/B must first be placed in network analyzer language before running the program.

- 8757C/D/E scalar network analyzer
- 436A power meter with an appropriate power sensor

- 10 dB attenuator
- Printer: (optional) for printouts; the printer must be capable of performing a graphics dump.
- Three GPIB cables (4 required if a printer is used): 10833A/B/C/D
GPIB is Agilent's hardware, software, documentation, and support for IEEE-488 and IEC-625 worldwide standards for interfacing instruments.

There are two separate program listings:

- Cal Factor Entry Program
- Detector Frequency Response Program

Cal Factor Entry Program

The Cal (calibration) Factor Entry program allows you to input the calibration factors listed on the power sensor. These cal factors are stored on a disc under a file name that contains the power sensor serial number. This allows the storage of cal factors for more than one power sensor. At the beginning of the frequency response program, you are asked to input the serial number of the power sensor. Only data for that power sensor is loaded into memory. Up to 50 cal factors can be stored in each file although less are shown on the power sensors. A two-dimensional array is created containing combinations of a frequency and its associated cal factor. This program need only be run once to store the file on disc.

Running the Cal Factor Entry Program

The following information will help you successfully run the Cal Factor Entry program.

- When prompted, enter the last few digits of the power sensor serial number (no more than 5 digits).
- Enter the frequency in GHz, *not* MHz.
- Enter both the frequency and the cal factor. For example, if the cal factor at 50 MHz is 99, enter .05,99 and press **ENTER**.
- Each entered frequency must be greater than the preceding frequency.
- Fractional percentages are allowed (Example: 98.5).
- If you make a mistake, back up the program by entering a negative frequency. Each negative input backs up one entry and each entry must be input again.

- When all cal factors are entered, enter 0,0 to exit. The program displays all of the entered points. Verify the accuracy of each.
- When you press **CONTINUE**, the controller will store the file on disc. Make sure the disc is not write protected.

Detector Frequency Response Program

The Detector Frequency Response program performs the Frequency Response Performance Test using the cal factors previously stored on the disc. Any number of frequency points, up to 401, may be chosen, although 101 points is more than enough. The cal factor used is interpolated between the two closest frequency points. The cal factor used for frequencies measured below the lowest entered frequency point of the array defaults to the cal factor of the lowest entered frequency point. A similar method is used at the high end.

To reduce the time required to copy these programs, the code listings represent a minimum configuration to perform the measurement. There are no error checking or convenience features. Be careful when entering data and when following the displayed instructions.


Running the Detector Frequency Response Program

The following information will help you successfully run the Detector Frequency Response program.

- If not testing from .01 to 26.5 GHz, change the START/STOP frequencies in lines 180/190 to the appropriate values.
- When prompted, enter the last few digits of the power sensor serial number (no more than 5 digits). Use the same power sensor serial number as in the Cal Factor Entry program.
- When prompted, enter the number of frequency points to be taken (no more than 401). The data points will be evenly spaced across the frequency range.
- When prompted, connect the calibrated power sensor to the attenuated output of the source. The program automatically zeros the meter.
- When prompted, remove the power sensor and connect the detector under test to the attenuated output of the source. Connect the input cable of the detector to the "A" input of the scalar network analyzer.
- The program will ask if the test should be performed in AC or DC mode. Normally DC mode is used for frequency response testing.

Note If using DC mode, zero the detector before measuring the DUT response.

A sample displayed output of the program is shown in Figure 9-1. If the displayed frequency response exceeds the limits of the graph, increase the value of "Scale" (variable controlling the scale size) in line 210.



pc427a_e

Figure 9-1. Typical Program Output

Measurement Setup

Connect the equipment as shown in Figure 4-3. Connect the GPIB cables from the controller to the GPIB inputs of both the power meter and scalar network analyzer. Ensure that a system configuration and detector zero have been performed.

The GPIB addresses are as follows:

- Network analyzer: 716
- Source: 19 (connected to the analyzer's System Interface Bus, *not* the controller's GPIB)
- Power meter: 713

A printer may be connected to the GPIB at address 701 if a printout is required. The printer must be capable of performing a graphics dump.

Example Programs

The following programs can be used to automate the power meter CAL factor entry and the Frequency Response Performance Test.

```
10      ! POWER METER CAL FACTOR ENTRY PROGRAM
20      ! Creates 2 dim array of freq (in MHz) vs cal factor
        ! (in percent)
30      ! Assigns a file name of "PM_XXXXX" (XXXXX is the pwr
        ! sensor serial #)
40      !
50      OPTION BASE 0
60      DIM Cal(1:50,1:2)
70      INTEGER I,Pm_points
80      !
90      GRAPHICS OFF
100     PRINTER IS 1
110     OFF KEY
120     OUTPUT 1;CHR$(12)
130     BEEP 300,.1
140     INPUT "ENTER THE POWER SENSOR SERIAL # (LAST 5
        DIGITS)",Serial
```

```

150  !
160  PRINT USING "3/,K,/" ; "ENTER BOTH FREQ AND CAL SEPARATED
    BY A COMMA."
170  PRINT "ENTERING A NEG FREQ WILL BACK UP."
180  PRINT USING "3/,K,/" ; " #      FREQ      CAL%"
190  !
200  FOR I=1 TO 50  ! 50 POINTS MAX
210  Entry:  !
220      BEEP 200,.02
230      DISP I;" Enter FREQ (in GHz) and CAL (in %). ""0,0
    EXITS""";
240      INPUT "",Freq,Pct
250      IF Freq=0 OR Pct=0 THEN GOTO Done
260      Freq=Freq*1000
270      IF Freq<0 THEN
280          I=I-1
290          BEEP 1000,.02
300          GOTO Entry
310      END IF
320      IF I>1 THEN
330          IF Freq<=Cal(I-1,1) THEN
340              DISP "FREQ MUST BE GREATER THAN LAST ENTRY... TRY
    AGAIN!"
350              BEEP 2000,.5
360              WAIT 2
370              GOTO Entry
380          END IF
390      END IF
400      Cal(I,1)=Freq
410      Cal(I,2)=Pct
420      PRINT USING "DD,3X,6D.D,5X,3D.D";I,Freq,Pct
430  NEXT I
440  !
450  Done:  ! DISPLAY ENTERED VALUES
460      Pm_points=I-1
470      OUTPUT 1;CHR$(12)
480      PRINT USING "K,/,K,/" ; "      ENTERED VALUES"," #      FREQ
    CAL%"
490      FOR I=1 TO Pm_points
500          PRINT USING "DD,3X,6D.D,5X,3D.D";I,Cal(I,1),Cal(I,2)
510      NEXT I

```

```

520 PRINT "Power Sensor serial number =";Serial
530 BEEP 400,.1
540 DISP "PRESS CONTINUE TO STORE DATA ON DISC."
550 PAUSE
560 !
570 DISP
580 ON ERROR GOSUB Error
590 ASSIGN @File TO "PM_"&VAL$(Serial)
600 OFF ERROR
610 OUTPUT @File;Pm_points,Cal(*)
620 ASSIGN @File TO *
630 BEEP 500,.02
640 DISP "CAL FACTORS HAVE BEEN STORED ON DISC."
650 CAT
660 STOP
670 !
680 ! *****
690 Error: ! ERROR ON ASSIGNMENT
700 !
710 IF ERRN=56 THEN ! NO EXISTING FILE
720     CREATE BDAT "PM_"&VAL$(Serial),4 ! CREATE FILE
730     RETURN
740 END IF
750 DISP "ERROR # ";ERRN;" FIX, THEN PRESS CONTINUE"
760 BEEP 2000,.3
770 PAUSE
780 DISP
790 RETURN
800 END

```

```

10  ! DETECTOR FLATNESS PROGRAM
20  ! Plots detector measurements vs power meter measurements.
30  !
40  OPTION BASE 0
50  COM INTEGER Cal_data_flg
60  COM /Cal_f/ Cal(1:50,1:2),INTEGER Pm_points
70  COM /Measure/
    Meas(1:401,1:3),Set_power,Start,Stop,Scale,INTEGER Points
80  COM /Hpib/ @Sna,@Source,@Pwr_mtr
90  !
100 GINIT
110 CALL Clear_screen
120 OFF KEY
130 ASSIGN @Sna TO 716      ! Scalar Network Analyzer address
140 ASSIGN @Source TO 717  ! Passthrough address to Source
150 ASSIGN @Pwr_mtr TO 713 ! Power Meter address
160 !
170 ! Change the below variables as needed for the detector under
    test.
180 Max_points=401
190 Start=10      ! Start frequency in MHz
200 Stop=26510   ! Stop frequency in MHz
210 Set_power=2  ! Power level set point of source in dBm
220 Scale=2      ! dB (+/- graph limits)
230 !
240 IF NOT Cal_data_flg THEN ! get cal data on power sensor
250     DISP "ENTER THE POWER SENSOR SERIAL # (LAST 5 DIGITS)";
260     BEEP 700,.1
270     INPUT Serial
280     DISP "LOADING CAL FACTORS"
290     ASSIGN @File TO "PM_"&VAL$(Serial)
300     ENTER @File;Pm_points,Cal(*) ! Load cal factors
310     ASSIGN @File TO *
320     Cal_data_flg=1
330     DISP
340 END IF
350 BEEP 500,.1
360 DISP "ENTER NUMBER OF POINTS TO BE MEASURED
    (";Max_points;"MAX)";
370 INPUT Points
380 Step_size=(Stop-Start)/(Points-1)

```

```

390  FOR I=1 TO Points
400      Meas(I,1)=Start+Step_size*(I-1)
410  NEXT I
420  REMOTE 7
430  OUTPUT @Sna;"IP PT19" ! preset system
440  CALL Set
450  LOOP
460      CALL Verify
470      BEEP 300,.1
480      DISP "          To repeat...press CONTINUE;  to re-cal...RUN"
490      PAUSE
500      !
510  END LOOP
520  END
530  !
540  ! ***** SUB PROGRAMS *****
550  !
560  SUB Corr_pwr(Freq,Power) ! uses cal factor to find actual
power read
570      COM /Cal_f/ Cal(*),INTEGER Pm_points
580      IF Freq<Cal(1,1) THEN ! freq too low, use first
value
590          Cal_factor=Cal(1,2)
600          GOTO Act_pwr
610      END IF
620      X=0
630      REPEAT
640          X=X+1
650          IF X>Pm_points THEN ! freq too high, use last value
660              Cal_factor=Cal(Pm_points,2)
670              GOTO Act_pwr
680          END IF
690          UNTIL Freq<Cal(X,1)
700          Frac=(Freq-Cal(X-1,1))/(Cal(X,1)-Cal(X-1,1))
710          Cal_factor=Frac*(Cal(X,2)-Cal(X-1,2))+Cal(X-1,2) ! cal
factor in %
720  Act_pwr:Power=Power-(10*LGT(Cal_factor/100))          ! actual
power read
730  SUBEND
740  !
750  SUB Set ! finds freq response of source using pwr mtr as ref.

```

```

760     COM /Cal_f/ Cal(*),INTEGER Pm_points
770     COM /Measure/ Meas(*),Set_power,Start,Stop,Scale,INTEGER
        Points
780     COM /Hpib/ @Sna,@Source,@Pwr_mtr
790     Clear_screen
800     OUTPUT @Sna;"MDO"
810     OUTPUT @Source;"MDO CW PL-60DB RFO" ! works for 8350 or
        8340/41
820     BEEP 400,.1
830     DISP "CONNECT POWER SENSOR TO ATTENUATORS"
840     PAUSE
850     !
860     CALL Zero_mtr
870     OUTPUT @Source;"PL";Set_power;"DB RF1 MDO SV1 AMO PM1" !
        for 8350/40/41
880     FOR I=1 TO Points
890         OUTPUT @Source;"CW";Meas(I,1);"MZ"
900         IF I=1 THEN WAIT 5
910         Read_pwr(Meas(I,2))
920         Corr_pwr(Meas(I,1),Meas(I,2))
930         DISP USING 940;"Freq =";Meas(I,1);"MHz      Power
            =";Meas(I,2);"dBm"
940         IMAGE    K,6D.D,K,3D.2D,K
950     NEXT I
960     DISP
970     Set_flg=1
980     SUBEND
990     !
1000    SUB Clear_screen ! clears alpha and graphics screen
1010    GRAPHICS OFF
1020    OUTPUT 1;CHR$(12)
1030    DISP
1040    SUBEND
1050    !
1060    SUB Read_pwr(Power) ! reads power from 436A power meter @
        100%
1070    COM /Hpib/ @Sna,@Source,@Pwr_mtr
1080    REPEAT
1090        WAIT .3
1100        OUTPUT @Pwr_mtr;"9D+T"
1110        ENTER @Pwr_mtr USING "B,B,X,K";Sts,Range,Power

```

```

1120     UNTIL Sts=80
1130 SUBEND
1140 !
1150 SUB Zero_mtr
1160     COM /Hpib/ @Sna,@Source,@Pwr_mtr
1170     DISP "STAND-BY... zeroing power meter"
1180     OUTPUT @Pwr_mtr;"A1+R"
1190     WAIT 7
1200     OUTPUT @Pwr_mtr;"Z1+R"
1210     OUTPUT @Pwr_mtr;"9D+V"
1220     WAIT 7
1230     DISP
1240 SUBEND
1250 !
1260 SUB Verify ! measures detector response
1270     COM /Cal_f/ Cal(*),INTEGER Pm_points
1280     COM /Measure/ Meas(*),Set_power,Start,Stop,Scale,INTEGER
        Points
1290     COM /Hpib/ @Sna,@Source,@Pwr_mtr
1300     INTEGER I
1310     Clear_screen
1320     OUTPUT @Sna;"C2 CO C1 IA ME FDO MD1 SWO AF1"
1330     BEEP 500,.1
1340     DISP "Select AC or DC mode. (A or D) Default = AC ";
1350     INPUT Answ$
1360     IF Answ$[1,1]="d" OR Answ$[1,1]="D" THEN
1370     OUTPUT @Sna;"DMO MDO"
1380     ELSE
1390     OUTPUT @Sna;"DM1 MD1"
1400     END IF
1410     BEEP 400,.01
1420     DISP "CONNECT DETECTOR TO ATTEN AND TO INPUT ""A"" "
1430     PAUSE
1440     !
1450     ALPHA OFF
1460     Graticule
1470     GRAPHICS ON
1480     PEN 1
1490     OUTPUT @Source;"PL";Set_power;"DB RF1"
1500     FOR I=1 TO Points
1510         OUTPUT @Source;"CW";Meas(I,1);"MZ"

```

```

1520         WAIT .3
1530         OUTPUT @Sna;"0V"
1540         ENTER @Sna;Meas(I,3)
1550         Db_err=Meas(I,3)-Meas(I,2)
1560         IF I=1 THEN MOVE Meas(I,1),Db_err
1570         PLOT Meas(I,1),Db_err
1580     NEXT I
1590 SUBEND
1600 !
1610 SUB Graticule ! generates graphics graticule
1620     COM /Measure/ Meas(*),Set_power,Start,Stop,Scale,INTEGER
        Points
1630     GCLEAR
1640     DEG
1650     LDIR 0
1660     !
1670     ! ** GRATICULE **
1680     X=Stop-Start
1690     Xmin=-.15*X
1700     Xmax=1.02*X
1710     Ymax=1.2*Scale
1720     Ymin=-1.5*Scale
1730     WINDOW Xmin,Xmax,Ymin,Ymax
1740     CLIP 0,X,-Scale,Scale
1750     FRAME
1760     LINE TYPE 3
1770     GRID 2000,Scale/5,0,0
1780     LINE TYPE 1
1790     AXES 1000,Scale/10,0,0,2,2
1800     CLIP OFF
1810     !
1820     ! ** X-AXIS LABEL **
1830     CSIZE 4.3
1840     FOR I=INT(Start/1000) TO Stop STEP 4000
1850         LORG 6
1860         MOVE I,-Scale
1870         LABEL I/1000
1880     NEXT I
1890     !
1900     ! ** Y-AXIS LABEL **
1910     CSIZE 4.

```

```
1920     FOR I=-Scale TO Scale STEP Scale/5
1930         LORG 8
1940         MOVE 0,I
1950         LABEL (INT(I*100+.5))/100
1960     NEXT I
1970     !
1980     ! ** LABELS **
1990     CSIZE 5
2000     LORG 4
2010     MOVE (Stop-Start)/2,Scale
2020     LABEL "ABSOLUTE dB ERROR vs FREQ."
2030     LORG 6
2040     MOVE (Stop-Start)/2,-Scale*1.1
2050     LABEL "FREQUENCY in GHz."
2060     LDIR 90
2070     MOVE Xmin,0
2080     LABEL "dB ERROR"
2090     SUBEND
```

— |

| —

— |

| —

Index

A

adjustments, 4-33

C

cable

lead identification, 2-3

characteristics, supplemental, 1-7

connector

mating, 2-3

contacting Agilent, iv

D

detector

connecting, 2-3

replacing, 7-2

E

electrostatic discharge (ESD), 2-1

cautions, 3-1

F

floor mat

use, 2-2

H

heel strap

use, 2-2

I

initial inspection, 2-1

input connector

cleaning, 8-2

gaging, 8-2

installation, 1-10

M

maintenance, 7-5

cleaning the connectors, 8-2

gaging the input connector, 8-2

mechanical inspection, 8-1

visual inspection, 8-2

mat

use, 2-2

module exchange program, 6-3

O

operating environment, 1-9

ordering parts, 6-1

P

packaging, 1-9

parts, ordering, 6-1

performance test record, 4-25-33

absolute power accuracy response,
4-29

alternate power accuracy tests,
4-31

dynamic accuracy response, 4-30

frequency response, 4-27

input connector mechanical

tolerances, 4-32

performance tests, 3-8

equipment required, 4-1

frequency response, 4-9-15

- frequency response:computing
 - maximum error, 4-14
- frequency response:detector
 - characterization, 4-14
- frequency response:equipment
 - required, 4-11
- frequency response:error analysis,
 - 4-10
- frequency response:procedure,
 - 4-13, 4-15
- frequency response:source
 - characterization, 4-14
- frequency response:specifications,
 - 4-13
- frequency response:system
 - configuration, 4-13
- power accuracy, 4-15-24
- power accuracy:absolute power in
 - DC mode, 4-17
- power accuracy:alternate
 - procedure, 4-21
- power accuracy:dynamic accuracy
 - in AC mode, 4-20
- power accuracy:equipment
 - required, 4-16
- power accuracy:procedure, 4-17
- return loss, 4-3
- return loss:failure, 4-9
- return loss:procedure, 4-3-8
- return loss:specifications, 4-3

power

- requirements, 2-3

product description, 1-1

program listing, 9-1

- automated cal factor entry program,
 - 9-2
- automated frequency response test,
 - 9-1

- example programs, 9-5

R

- repair procedures, 7-2
 - removing the covers, 7-2
 - replacing the cable assembly, 7-3
 - replacing the detector, 7-3
 - replacing the input connectors,
 - 7-5
- replaceable parts, 6-1
- replacing
 - detector, 7-2
- return loss
 - test description, 4-3

S

- service, 6-3
 - returning detector for service, 1-10
- specifications, 1-3-7
 - 85025A (including Option 001), 1-4
 - 85025B, 1-5
 - 85025D, 1-6
 - 85025E, 1-7
 - general, 1-3
- static-safe
 - practices, 2-2
 - work station, 2-2
- storage and shipment, 1-9
 - environment, 1-9

T

- table mat
 - use, 2-2

W

- wrist strap
 - use, 2-2



85025-90073

**Copyright Agilent Technologies, Inc.
2007**

Printed in Malaysia 11/2007

No. 85025-90073



Agilent Technologies