

Test Products International, Inc.

Calibration of ADF25, 25A (10/25/01)

1. How to open the case, please follow the steps.

1.1 Remove the battery lid.

1.2 You will see the two screws and loosen them.




Fig. 1

1.3 Near the BNC cable side, there are two protuberant mortises. Use fingers to push back and pull upward.




Fig. 2

1.4 Use fingers to press the bottom case inward, and pull it backward and upward.




Fig.3

1.5 Hold the bottom case and push back (to input lines direction).




Fig.4

1.6 You will see the two metal cases. There is one screw on the PCB. Loosen it, then the whole metal case can be taken out. There are two wires still connected with the upper case. Leaves them there.




Fig. 5

1.7 The metal case is illustrated as follows:





Fig. 6

2. Explanation of adjustment point. Please refer to Fig. 7.

- (1) SW1: Switch for selecting attenuation ratio.
- (2) SW2: Power switch.
- (3) VR2: Variable resistor for adjusting offset.
- (4) VR4: Variable resistor for adjusting offset.
- (5) VR1: Variable resistor for adjusting low frequency CMRR.
- (6) VR3: Variable resistor for adjusting accuracy.



3. Adjustment Procedure:

Please connect a 6VDC power adapter to CON1, and turn SW2 on to warm up for 10 minutes. Then follow the steps below.

3.1 Offset Adjustment:

- a. Referring to Fig. 8, connect both input lines together.
- b. Connect output BNC connector to a DMM.
- c. Put SW1 at 1/10 (or 1/20).
- d. Adjust VR2 to make output voltage approach to zero as small as possible.
- e. Put SW1 at 1/100 (or 1/200).
- f. Adjust VR4 to make output voltage approach to zero as small as possible.
- g. Repeat “c” to “g” steps until output voltage is always within $\pm 0.1\text{mV}$ at either “e” or “g” step.




Fig. 8

3.2 Low Frequency CMRR Adjustment

- Referring to Fig. 9, connect output BNC to the oscilloscope which has good grounding.
- Connect both input lines to a 220/200VAC, 60/50Hz power source.
- Put SW1 at 1/10 (or 1/20).
- Set the VOLTS/DIV switch on oscilloscope's panel to 1mV and the TIME/DIV switch to 5ms. Then a 60/50Hz sinusoid waveform should be displayed on the screen.
- Adjust VR1 to make the amplitude of displayed waveform as small as possible. Finally the amplitude should be less than 1.5mV.




Fig. 9

3.3 Square Wave Adjustment

- Referring to Fig. 10, connect output BNC to the oscilloscope has good grounding.
- Connect the red input line to a 20Vp-p, 10KHz square wave signal source and the black input line to ground.
- Put SW1 at 1/10 (or 1/20).
- Set the VOLTS/DIV switch on oscilloscope's panel to 0.1V for 1/10 (or 1/20) attenuation and the TIME/DIV switch to 20us.
- Adjust VC1 to make the displayed waveform become as Fig 11(b).

- f. Connect the black input line to 20Vp-p, 10KHz square wave signal source and the red input line to ground.
- g. Adjust VC2 to make the displayed waveform become as Fig. 11(b).




Fig. 10




Fig. 11

3.4 Repeat 3.2 and 3.3 adjustment until both adjustments are O.K.

3.5 High Frequency CMRR adjustment

- Referring to Fig. 12, connect output BNC to the oscilloscope has good grounding.
- Connect both input lines to a 155Vp-p, 20KHz, sine wave signal source.
- Put SW1 at 1/10 (1/20).
- Set VOLTS/DIV switch on oscilloscope's panel to 1mV and the TIME/DIV switch to 0.1ms.
- Adjust VC1 slightly to make the amplitude of displayed waveform as small as possible. Finally the amplitude should be less than 2mV.




Fig. 12

3.6 Accuracy Adjustment

- Referring to Fig. 13, connect output BNC to a DMM.
- Connect the red input line to the positive terminal of a DC voltage source (70VDC) and the black input line to the negative terminal source (70VDC).
- Put SW1 at 1/10 (or 1/20).

d. Adjust VR4 to make the output voltage equal to 6.9V (or 3.51V for 1/20).




Fig. 13