




DOC026.53.80202

MET ONE 3400: 3413, 3415, 3423, 3425, 3445

06/2013, Edition 5, Firmware version 4.08.XX

User Manual



Specifications	5
Instrument specifications	5
Sample measurement specifications	6
Battery specifications	6
General information	6
Safety information	7
Use of hazard information	7
Precautionary labels	7
Compliance	8
Country-specific approval for Wi-Fi devices	8
Certification	9
General product information	9
Installation	9
Unpack the instrument	9
Wiring safety information	10
Electrostatic discharge (ESD) considerations	10
Electrical connections	11
Install the batteries	12
Assemble the particle counter system	14
Install the printer paper	15
Connect RS485 communication	15
Network and communications	17
About network and communications setup	17
Setup for serial communication	17
Setup for Ethernet communication	17
Setup for wireless communication	17
Set wireless security	18
Particle counter navigation	18
Configuration	19
About configuration	19
About Basic and Advanced operation	19
Set the operation mode at initial startup	19
Change the operation mode	19
Configure the system	20
Set the time and date	20
Set the sleep mode and backlight timeout	20
Set the Alarm Reasons option	20
Set the Sample Comments option	21
Set the user interface language	21
Manage units and alarms	21
Manage audible alarm settings	22
Set the measurement units	22
Set the flow rate alarm values	22
Add an alarm reason	22
Edit an alarm reason	22

Table of Contents

Delete an alarm reason	22
Apply an alarm reason to a data record	23
Set the inert gas and altitude values	23
Set the particle count alert (beep function)	23
Change the relative humidity and temperature probe	23
Manage backup and restore settings	24
Make a backup of configurable settings	24
Restore settings from backup	24
Configurations	24
Copy a configuration	24
Install a configuration	24
Locations, areas and groups	25
Location management	25
Add a location	25
Edit a location	25
Configure new settings for the location	25
Copy settings from another location	26
Set location alarms	26
Remove a location	26
Change the order of locations	26
Area management	27
Add a new area	27
Edit an area	27
Remove an area	27
Change the order of areas	27
Group management	27
Add a group	27
Install a group	27
Delete a group	28
Add a location to a group	28
Sizes function	28
Data management	28
Store partial data	28
Manage the data buffer	28
Set the data buffer to rotate data	28
Set the data buffer size	29
Real-time PDF/CSV option	29
Turn on the PDF option	29
Save the count data to a folder	29
Manage users and permissions	30
Enable the user logon function	30
Log on as administrator	30
Change the password	30
Replace a forgotten password	30
Add a user	31
Assign groups to a user	31
Assign user access rights	31
Operation	32

Log on to the particle counter	32
Measure particle counts	32
Change the particle count location	32
See settings during the particle count	32
See historical data during the particle count	32
Use the filter scan probe	33
Manage sample batch identification	33
Enter or change a Batch ID	33
Disable a Batch ID	33
Set or clear the Batch ID	33
How to use the Test and Report Wizard	33
About standard sampling protocols	33
About reports	34
Set up the Test and Report Wizard	34
Start sample measurement with the Wizard	34
Use existing data	35
Report test results	35
How to use the Print Center	35
About the Print Center	35
Print records manually	35
Set automatic print functions	37
Review historical buffer data	37
Set the data filter	37
Clear the data buffer	37
Export data	37
About status values in exported data	38
Configure and enable the FTP function	39
Data transfer to the OPC server	39
User-initiated data upload	39
User-initiated data download	39
Automatic data download	39
Maintenance	40
Clean the instrument exterior	40
Set the count to zero	40
Update the instrument software	40
Charge the batteries in the particle counter	40
Battery recharge intervals	41
Calibrate the battery	41
Diagnostics and Troubleshooting	41
Factory settings	42
Parts and accessories	42
Parts for the 28.3 LPM counter (3413 and 3415)	42
Parts for the 50 LPM counter (3423 and 3425)	42
Parts for the 100 LPM counter (3445)	43
3400 series parts	43
Spare parts kit (2087919-01)	44

Specifications

Specifications are subject to change without notice.

Instrument specifications

Specification	Detail
Power requirement	Adapter (included in the ship kit): 100–240 VAC, 2.5 A, 50–60 Hz
	Instrument: 24 VDC, 75 W maximum
Installation category	I
Protection class	III
Pollution degree	2
Altitude	2000 m (6562 ft)
Light source	Long Life Laser™ diode with 10-year Mean Time To Failure (MTTF), Class 3B Laser, 810 to 852 nm, 50 mW maximum
Pump type	Air vacuum, rated for continuous use
Count display	Color ¼ VGA TFT touch screen
Interface	Windows CE®-based
Maximum count shown	9,999,999
Delay time	00:00:06 to 23:59:59
Sample and hold times	Sample: 00:00:01 to 23:59:59
	Hold: 00:00:00 to 23:59:59
Count alarms	1 to 9,999,999 counts
Data storage	50 to 5000 samples, scrollable on Historical Data review screen 3000 is the default value
Count cycles	Up to 100 while in automatic mode
Locations	Up to 999
Exhaust port	3/8-in. NPT thread
Outputs	Ethernet–10Base-T/100Base-TX
	RS485 Serial
	RS232 Serial
	Optional wireless–802.11 b/g compatible
	USB Client (Version 1.1)
	USB Host (Version 1.1)
Manifold	Supports A3432, 32-port manifold system (available on 1 CFM units only)
Enclosure material	Stainless steel
Weight without battery	3413 and 3415—7.55 kg (16.6 lb)
	3423 and 3425—8.33 kg (18.3 lb)
	3445—8.65 kg (19.0 lb)
Size (W x D x H)	31.8 x 25.4 x 20.3 cm (12.5 x 10 x 8 in.)

Specification	Detail
Environment, operation	0 to 40 °C (32 to 104 °F); 10 to 90% relative humidity, non-condensing
Environment, storage	–40 to 50 °C (–40 to 122 °F); 0 to 98% relative humidity, non-condensing

Sample measurement specifications

Sampling	
Number of size ranges	Standard 6, 8
Particle size ranges and standard channels	Models 3413 and 3423—0.3, 0.5, 1.0, 3.0, 5.0, 10.0 µm
	Models 3415 and 3425—0.5, 1.0, 2.0, 3.0, 5.0 and 10.0 or 25.0 µm
	Models 3445—0.5, 1.0, 2.0, 3.0, 5.0, 10.0 µm
Flow rate	Models 3413 and 3415—28.3 L/min (1.00 cfm) ± 5% (Default factory setting)
	Models 3423 and 3425—50 L/min (1.77 cfm) ± 5% (Default factory setting)
	Model 3445—100 L/min (3.53 cfm) ± 5% (Default factory setting)
Zero count	Conforms to JIS B9921. 1 count or less in 5 minutes, 95% confidence level
Coincidence loss	Models 3413 and 3415—10% at 20,000,000 particles/m ³ (566,570 particles/ft ³)
	Models 3423 and 3425—10% at 10,000,000 particles/m ³ (283,280 particles/ft ³)
	Model 3445—10% at 5,000,000 particles/m ³ (141,640 particles/ft ³)
Count efficiency	Models 3413 and 3423—50% ± 20 % for 0.3 µm, (100% ± 10% at 1.5 times the minimum sensitivity). Fully complies with ISO21501-4.
	Models 3415, 3425 and 3445—50% ± 20% for 0.5 µm, (100% ± 10% at 1.5 times the minimum sensitivity). Fully complies with ISO21501-4.

Battery specifications

Specification	Detail
Battery type	Lithium ion smart battery; can be charged, ejected and changed without disruption to the system.
Quantity included	One (two batteries are provided with the 3445)
Battery life during operation	Models 3413 and 3415—6 hours
	Models 3423 and 3425—7 hours
	Model 3445—3.5 hours
Battery recharge time	6.75 hours minimum, 10 hours maximum
Power	14.4 VDC, 6.6 Ah (2x)
Battery weight	0.66 kg (1.45 lb)

General information

In no event will the manufacturer be liable for direct, indirect, special, incidental or consequential damages resulting from any defect or omission in this manual. The manufacturer reserves the right to make changes in this manual and the products it describes at any time, without notice or obligation. Revised editions are found on the manufacturer's website.

Safety information

NOTICE

The manufacturer is not responsible for any damages due to misapplication or misuse of this product including, without limitation, direct, incidental and consequential damages, and disclaims such damages to the full extent permitted under applicable law. The user is solely responsible to identify critical application risks and install appropriate mechanisms to protect processes during a possible equipment malfunction.

Please read this entire manual before unpacking, setting up or operating this equipment. Pay attention to all danger and caution statements. Failure to do so could result in serious injury to the operator or damage to the equipment.

Make sure that the protection provided by this equipment is not impaired. Do not use or install this equipment in any manner other than that specified in this manual.

Use of hazard information

⚠ DANGER

Indicates a potentially or imminently hazardous situation which, if not avoided, will result in death or serious injury.

⚠ WARNING

Indicates a potentially or imminently hazardous situation which, if not avoided, could result in death or serious injury.

⚠ CAUTION

Indicates a potentially hazardous situation that may result in minor or moderate injury.

NOTICE

Indicates a situation which, if not avoided, may cause damage to the instrument. Information that requires special emphasis.

Precautionary labels

Read all labels and tags attached to the instrument. Personal injury or damage to the instrument could occur if not observed. A symbol on the instrument is referenced in the manual with a precautionary statement.

	This symbol, if noted on the instrument, references the instruction manual for operation and/or safety information.
	This symbol, when noted on a product enclosure or barrier, indicates that a risk of electrical shock and/or electrocution exists.
	Delicate internal electronic components can be damaged by static electricity, resulting in degraded performance or eventual failure.
	This symbol indicates a laser device is used in the equipment.

	This symbol identifies the location of a fuse or current limiting device.
	Electrical equipment marked with this symbol may not be disposed of in European public disposal systems after 12 August of 2005. In conformity with European local and national regulations (EU Directive 2002/96/EC), European electrical equipment users must now return old or end-of-life equipment to the Producer for disposal at no charge to the user. Note: For return for recycling, please contact the equipment producer or supplier for instructions on how to return end-of-life equipment, producer-supplied electrical accessories, and all auxiliary items for proper disposal.

Compliance

CLASS 1 LASER PRODUCT	This symbol indicates that the instrument is a Class 1 LASER product.
------------------------------	---

This product complies with IEC/EN 60825-1:2007 and 21 CFR 1040.10 except for deviations pursuant to Laser Notice No. 50, dated June 24, 2007. FDA accession number: 9020917.

This product is also CE compliant. Contact the manufacturer for complete compliance details.

Country-specific approval for Wi-Fi devices

⚠ CAUTION	
	Electromagnetic radiation hazard. Make sure that the antenna is kept at a minimum distance of 20 cm (7.9 in.) from all personnel in normal use. The antenna cannot be co-located or operated in conjunction with any other antenna or transmitters.

Products with the wireless option contain a modular RF Wi-Fi device that operates in the 2.4 GHz range.

- United States FCC ID: R68WIPORTG
- Canada IC ID: 3867A-WIPORTG

Country	ISO31662 letter code	Country	ISO31662 letter code
Austria	AT	Poland	PL
Belgium	BA	Portugal	PT
Denmark	DK	Spain	ES
Finland	FI	Sweden	SE
France	FR	United Kingdom	GB
Germany	DE	Iceland	IS
Greece	GR	Norway	NO
Hungary	HU	Switzerland	CH
Ireland	IE	Turkey	TR
Italy	IT	Netherlands	NL
Mexico	MX	—	—

Regulatory RF device approvals

- FCC: Approved as a Modular Device under a TCB Grant of Authorization. FCC ID: R68WIPORTG
- IC: Approved as a Modular Device under Certificat D'Acceptabilite' Technique C-REL ID : 3867A-WIPORTG

Opinion: Compliant under the R&TTE Directive 1999/5/EC to the essentials requirements of Article 3.2 according to the assessment procedures in Article 10(5) and Annex IV for (class-2 equipment) and marked as CE1177.

Certification

The device complies with Part 15 of the FCC Rules and Industry Canada license-exempt RSS standard(s). Operation is subject to the following conditions:

- 1. The equipment may not cause harmful interference.
- 2. The equipment must accept any interference received, including interference that may cause undesired operation.

Changes or modifications to this wireless communication equipment not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment. Any change to the equipment will void the Industry Canada certification and FCC grant.

General product information

This manual describes the use of the MET ONE 3400 Series Particle Counter. The MET ONE 3400 Series Particle Counter counts and measures the size of airborne particles in cleanroom environments. Refer to [Table 1](#).

Table 1 MET ONE 3400 Series particle counter model numbers

Model number	Flow rate		Minimum particle size channel (µm)
	L/min	Ft³/min	
3413	28.3	1	0.3
3415	28.3	1	0.5
3423	50	1.77	0.3
3425	50	1.77	0.5
3445	100	3.53	0.5

Installation




⚠ WARNING

Multiple hazards. Only qualified personnel must conduct the tasks described in this section of the document.

Unpack the instrument

Remove all items from the shipping container and inspect for damage. If any items are missing or damaged, contact the manufacturer. Refer to [Figure 1](#).

Figure 1 MET ONE 3400 components



1 3400 Series Particle Counter	8 Extension tube for isokinetic probe
2 Rechargeable battery (280-120-2024)	9 Zero count filter
3 AC-to-DC power supply (280-300-5000)	10 RS485 connector assembly
4 Power cord (US)	11 USB Flash drive
5 Power cord (EU)	12 Intake cleaning brush
6 Thermal paper rolls for printer (2x)	13 Wireless antenna for Wi-Fi
7 Isokinetic probe	14 Stylus for touchscreen interface

Wiring safety information

⚠ WARNING
Electrocution hazard. Make sure that there is easy access to the local power disconnect.

NOTICE

Always disconnect power to the instrument before electrical connections are made.

Obey all safety statements while connections are made to the instrument.

Electrostatic discharge (ESD) considerations

NOTICE
Potential Instrument Damage. Delicate internal electronic components can be damaged by static electricity, resulting in degraded performance or eventual failure.


Refer to the steps in this procedure to prevent ESD damage to the instrument:

- Touch an earth-grounded metal surface such as the chassis of an instrument, a metal conduit or pipe to discharge static electricity from the body.
- Avoid excessive movement. Transport static-sensitive components in anti-static containers or packages.
- Wear a wrist strap connected by a wire to earth ground.
- Work in a static-safe area with anti-static floor pads and work bench pads.

Electrical connections


Connect probes, external power, cables and USB devices as shown in [Figure 2](#) and [Figure 3](#).

Figure 2 Back view



1 Exhaust port	4 Manifold controller connector or standard RS232 port (available on 1 CFM units only)
2 Serial communications RS485 connector	5 Battery ports
3 Auxiliary I/O port for the filter scan probe	6 Supplemental feet

Figure 3 Front and side view







1 Sample intake nozzle	8 USB host connector
2 Touchscreen	9 USB client connector
3 Power button	10 Relative humidity and temperature probe connector
4 Battery status indicator	11 Air velocity probe connector
5 Ethernet connector	12 Handle
6 Handle	13 Wireless antenna connector
7 Printer	14 Power connector

Note: For best results, use USB flash drives supplied by the manufacturer. Contact Customer Service for additional support at 800.866.7889 or +1.541.472.6500.

Install the batteries

⚠ WARNING


Explosion hazard. To avoid fire and/or explosion, use only the battery type and power supply/charger specified by the manufacturer. For part numbers, refer to [Figure 1](#) on page 10.



Assemble the particle counter system

Figure 4 shows the setup of the particle counter system.


Figure 4 Particle counter assembly



Install the printer paper





To prevent damage to the printer, always operate the particle counter with the recommended thermal paper installed in the printer. If the particle counter must be used without paper, set the print mode to "None". To install a roll of thermal printing paper, refer to [Figure 5](#).

Figure 5 Printer paper installation



Connect RS485 communication

To connect for RS485 communication, refer to steps 1 through 4 of this procedure.



Network and communications

NOTICE

Only qualified personnel should perform the tasks described in this section.

About network and communications setup



This section shows the setup for:

- Serial communications
- Ethernet network communication
- Wireless (Wi-Fi) communication
- Wireless security

Setup for serial communication

NOTICE

RS232 communication and manifold support cannot be used together because they use the same serial port.

1. On the Counter Navigation screen, push **NETWORK**.
2. Select the Serial tab.
3. Configure these options:
 - Select FX or Modbus RTU.
 - Set the ID for FX or Modbus.
 - Set the Baud rate.
 - Select RS232 or RS485.
 - Activate Manifold Support as needed (RS485 only).
 - Activate Auto Increment Port as needed

Setup for Ethernet communication

1. On the Counter Navigation screen, push **NETWORK**.
2. Select the Ethernet tab.
3. Configure these options:
 - Select FX or Modbus protocol.
 - Enter the Modbus port number.
 - Enter the Configuration port number.
 - Enter the IP address or select DHCP to let the network assign the IP address.
 - Enter the subnet mask address.
 - Enter the gateway address.

Setup for wireless communication

1. On the Counter Navigation screen, push **NETWORK**.
2. Select the Wi-Fi Configuration tab.
3. Configure these options:
 - Enter the Network name.
 - Select the data rate.
 - Select the network type: Ad hoc or infrastructure.
 - Enable auto fallback as needed.

- Select the channel (Ad Hoc mode only).
- Check Enable Radio if needed.
- Select the country.

Set wireless security

1. On the Counter Navigation screen, push **NETWORK**.
2. Select the Wi-Fi Security tab.
3. Configure these options:
 - Select the security type.
 - Select the encryption type.
 - Select the authentication type.
 - Select the key type: Hex or Passphrase.
 - Enter the key in the Key field and in the Retype Key field.

Particle counter navigation

The functions of the particle counter are accessed from the Counter Navigation screen. [Table 2](#) shows the functions that are accessible through the navigation screen.

Table 2 Screen icon descriptions

Icon	Function	Description
	Sample	Measure particle counts. Refer to Measure particle counts on page 32.
	Historical	Review measurement results in the buffer; print, export or filter data. Refer to Review historical buffer data on page 37.
	Export	Output file as comma separated value (CSV), tab separated, or PortAll files. Refer to Export data on page 37.
	Printer	Print sample data as hard-copy. Refer to How to use the Print Center on page 35.
	Locations	Add/edit/remove areas; copy location settings, edit locations settings; edit alarms for specific locations. Refer to Location management on page 25.
	Group	Load/add/edit a group; delete a group. Refer to Group management on page 27.
	System	Time/Date; Sleep time/backlight timeout; set logon requirements; set sounds for alarms; manage users; set the units for flow rates; manage the data buffer. Refer to the CD for more information. Refer to Configure the system on page 20.
	Sizes	Add/edit/delete a size (optional). Refer to Sizes function on page 28.

Table 2 Screen icon descriptions (continued)

Icon	Function	Description
	Test wizard	Test and report wizard for ISO, EU-GMP, FS or BS classification compliance. Refer to Set up the Test and Report Wizard on page 34.
	Return	Return to the previous screen or menu.

Configuration

About configuration

This section describes tasks that are usually done at the initial commissioning stage. Other tasks are done as updates to the particle counter are needed.

Note: If user passwords are set, the user must be logged in as an administrator to configure the particle counter.

About Basic and Advanced operation

At startup, the user can set the operation mode to Basic or Advanced settings.

Use the Basic mode to:

- Do all particle measurement functions
- Save run data in the buffer

Use the Advanced mode to:

- Export dialogs
- Measure air velocity
- Set analog alarms
- Set standards wizards and reports
- Backup, restore and restart
- Copy a configuration
- Set the network protocol (serial, Ethernet, Wi-Fi)
- Manage areas and locations
- Manage user counts and security
- Set user-configurable particle sizes
- Filter data

Set the operation mode at initial startup

To set the operation mode:

1. Power on the instrument.
2. While the operation mode selection screen shows, do one of the actions that follow:
 - Select **BASIC** or **ADVANCED**.
 - Wait 10 seconds. The instrument will start in the selected operation mode.

Change the operation mode

Only permissions in the Advance Operation mode allow a user to change the operation mode. To change operation mode from Advanced to Basic mode:

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Custom** tab.

3. Select **BASIC MODE**.
4. Push **RETURN** to restart the instrument.

Configure the system



System settings control how data is measured and stored, user permissions and other system-wide parameters. System settings can be configured as part of instrument commissioning, or changed later for different applications.

Set the time and date

Time and date will need to be set at initial commissioning.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Basic** tab.
3. Select the Time field. Use the numeric keypad to enter the current time in the HH:MM:SS format, then confirm.
4. Select the Date field. Use the numeric keypad to enter the current date in the YYYY-MM-DD format, then confirm.

Set the sleep mode and backlight timeout

Sleep mode and backlight time are active during battery use. When the instrument is connected to AC power, sleep mode and backlight are not active.

During sleep mode the instrument goes into hibernation after a period of inactivity to conserve power. All subsystems are shut down. The time value for sleep mode is in minutes.

The backlight setting turns off the LCD backlight after a period of inactivity to conserve power. The time value for the backlight setting is in seconds.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Basic** tab.
3. Select the Sleep Time field. Enter the length of idle time before sleep mode begins, then confirm. The range is 1 to 30 minutes.

Note: Enter 0 to disable sleep mode.

4. On the **Basic** tab, select the Backlight Timeout field. Enter the length of idle time before the user interface backlight turns off, then confirm. The range is 5 to 300 seconds.

Note: Enter a value of less than 5 seconds to disable backlight timeout.

To bring the instrument out of backlight timeout, use a finger or a stylus to make the user interface touchscreen active.

Set the Alarm Reasons option

To capture data about why alarms occur, the Alarm Reason option is set as Required or Optional. In both Required and Optional mode, the user is prompted to select from preset alarm reasons for the alarm event. All users can enter alarm reasons. Only administrators, factory-level users or users with System Settings permissions can disable or change the Alarm Reasons option.

To enable or disable Alarm reasons:

1. On the Counter Navigation screen, push **SYSTEM**.
2. On the **Units and Alarms** tab, push .
3. On the Alarm Reasons screen, select **REQUIRED**, **OPTIONAL** or **DISABLED**. Refer to [Table 3](#) for more information.

Note: The Alarm Reasons option is set to Disabled by default.

Table 3 Alarm Reasons options

Option	Description
Required	After doing a user-initiated sample cycle, the user is required to enter information about all alarms that happen during the cycle. Alarm reasons must be defined in the Alarm Reasons screen. The user cannot exit the Alarm History screen until a reason has been entered for all alarms.
Optional	After doing a user-initiated sample cycle, the user is prompted to enter information about all alarms that happen during the cycle. Alarm reasons must be defined in the Alarm Reasons screen.
Disabled	The user is not prompted to enter information about alarms.

Set the Sample Comments option

To record user comments made during routine sampling, change the Sample Comments setting. Only administrators, factory-level users or users with System Settings permissions can change the Sample Comments setting.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select **Sample Comments**.
3. Select an option in the field at the bottom of the screen.

Option	Description
Required	After each user-initiated sample cycle, the user must select one of the preset sample comments from the sample comment list or enter a new comment.
Required on Alarm	After each user-initiated sample cycle, the user must select one of the preset sample comments from the sample comment list or enter a new comment if an alarm(s) occurred during the sample cycle.
Optional	After each user-initiated sample cycle, the user is asked if they would like to enter a comment for the sample. The user can then select one of the preset sample comments from the sample comment list or enter a new comment.
Disabled (default)	The user is not asked if they would like to enter a comment for the sample after each user-initiated sample cycle.

4. To add comments to the comment list so they can be quickly selected:
 - a. Select **Add**.
 - b. Use the keypad to enter a comment about a sample run.
 - c. Push **Enter**.

Set the user interface language

The user interface language can be changed at any time. A language change will require a system restart.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Options and Accessories** tab.
3. Select the Language field to show language options.
4. Select a language. Push  then **OK** to restart the instrument.

Manage units and alarms

Units and alarms are controlled on the **Units and Alarms** tab. Units are set for temperature, air velocity, relative humidity and flow rate. The high and low parameter for the flow rate alarm are also enabled and set in this screen. The instrument can be set to work with the Vaisala HMP probe. Preset alarm reasons are accessible from this screen.

Manage audible alarm settings

The **Sounds** tab lets the user select the sound and volume to confirm user interface actions. Sounds that are used for other alarms (stop errors, limit alarms and warnings) are selected in this tab.

1. On the Counter Navigation Screen, push **SYSTEM**.
2. Select the **Sounds** tab.
3. Select the User Feedback field.
4. Select a sound from the list of available notification sounds.
5. Select sounds for stop error, alarm limit and warnings from the list of available notification sounds.
6. Use the slider to set the volume.

Set the measurement units

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Units and Alarms** tab.
3. Select the Temperature field. Select **DEG C** (Celsius) or **DEG F** (Fahrenheit).
4. Select the Air Velocity Field. Select **MM/SEC** or **FT/MIN**.
5. Select the Flow Rate field. Select **LPM** or **CPM**.

Set the flow rate alarm values

Contact technical support before the flow rate alarm is changed.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Units and Alarms** tab.
3. Select **ENABLED**.
4. Select the High field. Enter a value between 5 and 20.
5. Select the Low field. Enter a value between 5 and 20.

Add an alarm reason

Alarm Reasons describe what conditions have caused an alarm. If a needed Alarm Reason is not available in the list, it can be added and then applied to future data records.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Units and Alarms** tab.
3. Push .
4. Push **ADD**. Enter a reason (up to 29 characters).
5. Push **ENTER** to confirm.

Edit an alarm reason

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Units and Alarms** tab.
3. Push .
4. Select a reason from the list.
5. Push **EDIT**. Change the text string as needed.
6. Push **ENTER** to confirm.

Delete an alarm reason

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Units and Alarms** tab.

3. Push .
4. Select a reason from the list.
5. Push **REMOVE**.
6. Push **ENTER** to confirm.

Apply an alarm reason to a data record

Prerequisites:

- Alarm reasons must be set to **ENABLED** or **OPTIONAL**. Refer to [Set the Alarm Reasons option](#) on page 20 for more information.
- Alarm reasons must be entered in the alarm reasons list before they can be applied to a data record. Refer to [Add an alarm reason](#) on page 22 for more information.

An alarm reason can be applied to any data record that has an alarm. Alarm reasons show in the Historical data screen and on printouts. Alarm reasons are also included in FTP transfers and all USB flash drive exports.

1. On the Counter Navigation screen, push **HISTORICAL**.
2. Select the Data Buffer field. Enter the number of the record with the alarm, or go to the record with **+** and **-** keys.
3. Select the yellow text below Counts to show the Alarm Reasons list.
Note: The yellow text shows the type of alarm that has been recorded.
4. Select an alarm reason from the list. Push **ENTER**.

Set the inert gas and altitude values

After factory calibration, an inert gas can be selected. The particle counter applies a correction factor to the flow rate based on the inert gas selection.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Custom** tab.
3. Select the Gas field, then select the gas to be tested.
Note: Only the gases that the instrument has been calibrated with will show in the list.
4. Select the Altitude field.
5. Select the altitude of the measurement location. The altitude values are shown in feet.

Set the particle count alert (beep function)

The instrument can be set to play a sound when an interval value is counted on a specified channel. For example, when the interval value is 100, the counter will beep each time it counts 100 particles.

1. On the Counter Navigation screen, push **SAMPLE**.
2. Select the **Settings** tab on the right side of the screen.
3. Select Quick Settings.
4. Push **YES** to edit settings for the default group.
5. Select the Run Mode field, then select **BEEP**.

Change the relative humidity and temperature probe

The standard relative humidity and temperature probe is used in most applications. If an HMP probe is needed, it can be selected on the **Units and Alarms** tab.

The scale of the standard probe and the HMP probe are different. When measurements are different than expected, make sure that the HMP probe is selected or deselected as needed.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Units and Alarms** tab.
3. Select the **HMP RH/T** check box to activate or deactivate the HMP probe.

Manage backup and restore settings

Users with administrator or System Settings permissions can make a backup of configurable settings, and restore the backup settings as needed.

Make a backup of configurable settings

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Options and Accessories** tab.
3. Push **BACK-UP**.

An electronic copy of the current configuration is stored in instrument memory. This version can be restored with the Restore function.

Restore settings from backup

Use the Restore function when settings become corrupt, incorrectly modified or when the software performs abnormally.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Options and Accessories** tab.
3. Push **RESTORE**.
4. Push **RESTART**.

Configurations

Specific configurations can be saved as a backup and/or copied as necessary between particle counters.

Copy a configuration

To copy a configuration:

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Configuration** tab.
3. Insert a USB drive into the USB host connector. Refer to [Figure 3](#) on page 12.
4. Push **COPY CONFIGURATION TO USB**.
5. A confirmation message will show. Push **OK**.
6. Remove the USB drive.

Install a configuration

To install a configuration from a USB drive:

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Configuration** tab.
3. Insert the USB drive with the configuration data into the USB host connector. Refer to [Figure 3](#) on page 12.
4. Push **READ CONFIGURATION FROM USB**.
5. A confirmation message will show.
6. Remove the USB drive.
7. Push **OK** to restart the instrument and load the new configuration.

Locations, areas and groups



A **location** defines a space, such as a work bench, that is identified for sample testing. To add, change or remove a location, refer to [Location management](#) on page 25.

An **area** is a group of defined locations that are geographically co-located. For example, Cleanroom A is an area. The work benches inside Cleanroom A may be defined as locations. To add, change or remove an area, refer to [Area management](#) on page 27.

Groups are locations that have common sampling attributes. For example, all locations that are tested every month may form one group. Locations in a group do not have to be co-located. To add, change or remove a group, refer to [Group management](#) on page 27.

Location management

Add a location

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an existing area, or if needed, create a new area and then select it. Push **ADD LOCATION**.
3. In the Add Sample Location screen:
 - Select the Location Name field. Enter the location name. Push **ENTER** to confirm.
 - Select the Location ID field to specify a numerical ID for the location. The numerical ID is unique for each location and must be in the range of 000 – 999.

Note: Use the **ALT** key to access special characters.

Edit a location

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an existing area.
3. Select **Edit Location**.
4. In the Add Sample Location screen:
 - Select the Location Name field. Enter the location name. Push **ENTER** to confirm.
 - Select the Location ID field to specify a numerical ID for the location.

Note: Use the **ALT** key to access special characters.

Configure new settings for the location

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an area.
3. Select a location within the area.
 - Push **ADD LOCATION** to configure a new location.
 - Push **EDIT LOCATION** to change the configuration of a current location.
4. In the Add Sample Location screen, push **SETTINGS**.
5. Select the **General** tab.
 - Set the count cycles and the count mode.
 - Select the run mode.
 - Set the count display option.
6. Select the **Timing** tab.
 - Set the duration for each sample.

- Set the sample hold time between count cycles.
- Set the sample delay time to allow delay before the sample test begins.

7. Push **RETURN** to go back to the Add Sample Location screen.

Copy settings from another location

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select the area that includes the destination location.
3. Select the location where the copied settings should go. Push .
4. Push **YES** to confirm or push **NO** to cancel.

Set location alarms

The particle counter allows different alarm settings for individual locations.

Note: Use this feature when the Use Location Settings is selected in Group Settings.

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an area.
3. Select a location within the area. Push **ADD LOCATION** or **EDIT LOCATION** as needed.
4. In the Sample Location Setup screen, push **ALARMS**. The Alarm Settings screen has two tabs to configure.
5. In the **Count** tab:
 - Edit the particle size
 - Edit particle concentration limits
6. In the **Environment** tab:
 - Enable the temperature alarm and set temperature limits.
 - Enable the relative humidity alarm and set relative humidity limits.
 - Enable the air velocity alarm and set air velocity limits.

Note: These settings are only valid with the specific environmental probe attached to the instrument.

7. Push **RETURN** to confirm and go back to the Add Sample Location screen.

Remove a location

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an area.
3. Select a location. Push **REMOVE LOCATION**.
4. Push **YES** to delete the location or push **NO** to cancel.

Change the order of locations

Change the listed order of locations to change the sample order during testing.

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an area.
3. Select a location. Push the **UP** or **DOWN** arrow to change the position of the location in the list.
4. Continue to select and move locations to create the needed order for samples.

Note: Arrow keys can also be used to move a location to a different area.

Area management

Add a new area

Use up to 15 alphanumeric characters to name an area.

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, push **ADD AREA**.
3. Enter the area name and confirm.

*Note: Use the **ALT** key to access special characters.*

Edit an area

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an existing area.
3. Select **EDIT AREA**. Confirm.

*Note: Use the **ALT** key to access special characters.*

Remove an area

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an area.
3. Push **REMOVE AREA**.
4. Push **YES** to delete the area or push **NO** to cancel.

Change the order of areas

Change the listed order of areas for ease of selection. Area order does not affect sample order.

1. On the Counter Navigation screen, push **LOCATIONS**.
2. On the Area/Location Setup screen, select an area.
3. Push the **UP** or **DOWN** arrow to change the position of the area in the list.
4. Continue to select and move areas to create the needed order.

Group management



A group is a collection of sample parameters and settings that can include locations. The locations in a group do not need to be geographically co-located.

Push the **±** button on the sample screen to go to the next location in the active group.

Add a group

1. On the Counter Navigation screen, push **GROUP**.
2. In the Group Settings Management window, select **<NEW>**. Enter the name of the group and confirm.

*Note: Use the **ALT** key to access special characters.*

3. Push **SAVE**.

Install a group

To enable a group, use the Load Group function.

1. On the Counter Navigation screen, push **GROUP**.
2. In the Group Settings Management window, select a group.
3. Push **LOAD**.

Delete a group

1. On the Counter Navigation screen, push **GROUP**.
2. In the Group Settings Management window, select a group.
3. Push **DELETE**.

Add a location to a group

1. On the Counter Navigation screen, push **GROUP**.
2. In the Group Settings Management window, select a group.
3. Push **SETTINGS**.
4. Push **ADD LOCATION**.
5. Navigate to the needed location and select it, then push **OK**.

Note: Push the **UP** or **DOWN** arrow to change the order of locations in the Group Settings screen. Refer to [Change the order of locations](#) on page 26 for more information.

Sizes function

NOTICE

If a selected channel has not been calibrated, particle counts will be based on interpolation. A maximum of $\pm 15\%$ size error and a $\pm 30\%$ counter error can occur when samples are taken at an interpolated size point.



The Sizes function is a separate option that must be enabled by the manufacturer. The icon will not show on the Counter Navigation screen if this option has not been purchased. Customized channel sizes can be selected at time of order.

With the Sizes function enabled, the particle counter can be configured for the particle sizes that follow*:

- 0.3 to 10- μm range: 0.3, 0.4, 0.5, 0.7, 0.8, 1.0, 2.0, 3.0, 4.0, 5.0, 7.0 and 10.0
- 0.5 to 25- μm range: 0.5, 0.6, 0.7, 0.8, 1.0, 2.0, 3.0, 5.0, 7.0, 10.0, 15.0 and 25.0

Note: Channel sizes 0.3 μm and 25.0 μm cannot be configured together.

Data management

Store partial data

Data can be collected from sample measurements that have been stopped by a user or by a flow error.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Custom** tab.
3. Select the **STORE PARTIAL DATA** checkbox.

Manage the data buffer

The data buffer allows the data to be preserved or overwritten. In addition, the size of the buffer can be changed and data can be cleared from the buffer.

Set the data buffer to rotate data

The default buffer setting is fixed. In this setting, no new data can be added to the buffer when it is full. When the buffer is set to rotate, new data can be added to a full buffer, while the oldest data record is deleted.

* Assumes 8-channel option is enabled

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Custom** tab.
3. Select **ROTATING BUFFER**.

Set the data buffer size

A change to the buffer size causes all current buffer data to be lost and unrecoverable.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Options and Accessories** tab.
3. Select the Data Buffer Size field. Enter a value between 50 and 5000.
4. Push **ENTER** to confirm.
5. Push **YES** to clear the data buffer.

Real-time PDF/CSV option

Use this option to send the count data to a PDF file and if necessary to a CSV file. Secure PDF files are generated in real time or when sample data is selected from the instrument data buffer using the Export function.

Turn on the PDF option

1. Push **SYSTEM**.
2. Select the **PDF** tab.
3. Enter the unlock code and push **ENTER**.
4. Select the options:

Option	Description
PDF on USB Detect	Automatically saves the real-time count data to a PDF file when the particle counter finds a USB memory stick in the USB port.
New Document On	Sets the criteria to start a new PDF file. New PDFs are made when an area or location changes or when the date changes.
Export Averages	Includes the average count of a set number of runs and the count data for each run to a PDF file.
Export to Excel	When a PDF file is made, the count data is also transmitted in .csv (comma separated value) spreadsheet format.
Warn when USB drive is X% full	Sets a notification that the USB memory stick is 50–95% full.

Save the count data to a folder

Count data can be saved to a file with a filename that includes the date and time the file is created, or a unique identifier.

1. At the particle counter, make a folder to hold the count data. Select Date or enter a name for the folder.
2. Select **Configure File Name**.
3. Use up to seven criteria options to make a file name. Options include: area, date, time, location, user name, serial number or text.
When the file name is configured, the PDF is automatically saved to the folder.
4. Select **<None>** to remove a previous entry.

Manage users and permissions



Passwords allow the system administrator to restrict access to the instrument settings. When passwords are enabled, there are two levels of access:

- Administrator—Access to all settings on the instrument except for service (factory) access
- Operator—Access to review historical data and read current measurement values in the Diagnostics section. Operator access can also print historical or diagnostics data.

If password protection is not enabled, all users can access all functional settings of the instrument.

Note: Factory settings are never accessible without a password.

Enable the user logon function

To enable user and password login:

1. Push **SYSTEM**.
2. Select the **Basic** Tab.
3. Select **User Logon Required**.

Log on as administrator

1. On the Counter Navigation or Sample screen, push  to log off.
2. Select the User Name field. Enter "Admin". Push **ENTER**.
3. Select the Password field. Enter the default password "123456". Push **ENTER**.
4. Push **OK**.

Note: To maintain system security, change the default administrator password. Refer to [Change the password](#) on page 30.

Change the password

The password can be changed at the logon screen.

1. From the Counter Navigation screen, push Factory.
2. Push **CHANGE PASSWORD**.
3. Enter the username and push **ENTER**.
4. Enter the old password and push **ENTER**.
5. Select the New Password field. Enter a new password and push **ENTER**.
6. Select the Confirm Password field. Enter the new password again. Push **ENTER**.
7. Push **OK**.

Note: To replace a forgotten password, contact the manufacturer with:

- The counter serial number
- The current date setting in the format MMDDYYYY.

Replace a forgotten password

If the System Administrator password is forgotten, contact technical support to get a new password. Technical support requires the information that follows to supply a new password:

- Instrument serial number
- Current date setting in the format MMDDYYYY, where MM is the two digit month, DD is the two digit day, and YYYY is the four digit year

Add a user

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Users** tab.
3. Push **ADD**.
4. Select the User Name field and enter a user name.
5. Select the Password field and enter a password.
6. Select the Confirm Password field and enter the same password.
7. Select the access level (Admin or Operator).

Assign groups to a user

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Users** tab.
3. Select a user from the list.
4. Push **EDIT**.
5. Select the **Groups** tab.
6. Push **ADD**.
7. In the Select Group screen, select a group from the list.
8. Push **OK**.

Assign user access rights

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Users** tab.
3. Push **EDIT**.
4. Select the **Access Rights** tab.
5. Select or deselect access rights as necessary.

Option	Description
Sampling	The user can start a new count cycle.
Report Wizard	The user can access the report wizard to do sample protocols based on ISO 14644-1, Federal Standard 209E, British Standard 5295, EU Annex I or based on averages.
Historical	The user can search existing data records in the buffer.
Export	The user can export data to the USB stick based on the contents of the buffer or the filtered contents.
Printer	The user can use the different printer functions.
Group Settings	The user can set up or modify group settings.
Area/Location	The user can set up or modify area and/or location settings.
Group Select	The user can select the active group.
Group Admin	The user can add, edit or remove groups and can select the active group.
System Admin	The user can add or remove users, set date/time or any other function found under the System Settings menus.
Diagnostics	The user can see and print current diagnostic information.

Option	Description
Network	The user can access, see and change the network settings.
User Upload	The user can manually trigger the electronic transfer of count records through Ethernet (wired or wireless).

6. Push **OK**.

Operation

Log on to the particle counter



Prerequisites

- Start the system.
 - Enable password protection. Refer to [Enable the user logon function](#) on page 30.
1. Activate the backlight with a finger or stylus if needed.
 2. Push the **LOCK** icon to log out a previous user. Push **LOCK** again to see the logon screen.
 3. Enter the user name and password. Confirm.

Note: Push the **ALT** key to access special characters.

Measure particle counts



After a complete particle count measurement, the number of particles measured will show on the screen and be stored as data. Other configured parameters, such as relative humidity, temperature and air velocity will be shown and stored in data.

1. Remove the protective cap from the inlet tube on the counter.
2. On the Counter Navigation screen, push **SAMPLE**.
3. To start the particle count, push the **RUN** button. The **RUN** button will change to a **STOP** button while the count is measured.

Note: Push **STOP** to end the test before the count is complete. Incomplete particle count data will not be stored or printed.

4. When the count measurement is complete, the test will stop automatically.

Change the particle count location

There are two methods to change the location for a particle count.

- On the Sample screen, push the location name. Select the new location name and confirm.
- On the Sample screen, push the **PLUS** button to increment the location, or push the **MINUS** button to decrement the location.

See settings during the particle count

Location and group settings can be seen at any time during the particle count cycle.

- On the Sample screen, select the Settings tab on the right side of the screen.

See historical data during the particle count

Historical sample data can be seen at any time during the particle count cycle.

1. On the Sample screen, select the **ARROW** button.
2. Select the **HISTORICAL DATA** icon to see the data.

Use the filter scan probe

NOTICE

The filter scan probe function applies to 1 CFM and 50 LPM units only.

1. On the Counter Navigation screen, push **SAMPLE**.
2. In the Test screen, push the **FILTER** icon.
3. To start the test, push **START FILTER PROBE TEST**.
4. To end the test, push **STOP FILTER PROBE TEST**.
5. Push **PRINTER** to generate a brief report of the last completed test.

Manage sample batch identification

Batch IDs are used to label different test runs. Labels can be text or numbers. The Batch ID is shown:

- On the main sample screen in the settings
- In the System Messages of the Historical screen
- In printed reports, FTP exports and all USB exports

Enter or change a Batch ID

To enter or change a Batch ID:

1. On the Counter Navigation screen, push **SAMPLE**.
2. Select the **Settings** tab.
3. Select **BATCH ID**.
4. Enter the Batch ID as a text, numerical or alphanumeric string. The Batch ID can use up to 29 characters.
5. Push **ENTER**.

Disable a Batch ID

Disable the Batch ID to stop it from showing in data reports.

1. On the Counter Navigation screen, push **SAMPLE**.
2. Select the **Settings** tab.
3. Select **BATCH ID**.
4. Delete the Batch ID text.
5. Push **ENTER**. The default Batch ID value will show, but will not be reported.

Set or clear the Batch ID

At the start of a sample test push **BATCH ID**. The user can select:

- Continue—Keep the Batch ID the same.
- Clear ID—No Batch ID is recorded with the sample cycle.
- Edit ID—Change the Batch ID for the current sample cycle.

The Batch ID can also be cleared when the sample cycle is complete.

How to use the Test and Report Wizard

About standard sampling protocols



Use the Test and Report Wizard for groups, areas and locations that require sample measurements and reports based on specific standards and regulatory guidelines. The 3400-series

particle counters include sample strategies based on common international protocols such as ISO 14644-1, FS 209E, BS 5295 and EU GMP Annex 1.

The wizard steps through seven data entry points:

- Selection of the type of standard or regulatory guideline: EU GMP, ISO 14644-1, FS209E, BS5295 or Averages (user-defined test protocols)
- Selection of the targeted class for the qualification of the room
- Occupancy state: As built, at rest, in operation
- Particle size or sizes to sample
- Unidirectional or non-directional air flow (for FS209E)
- Area of the room in square meters (m²)
- Samples per location. A minimum value is set based on room size and standard. This value can be increased as needed.

About reports

ISO 14644-1, FS 209E and BS 5295 specify calculations for airborne particle counter count data. These documents establish the definitions for level of cleanliness in cleanrooms and clean zones based on specified concentrations of airborne particles. The printed reports provide the data to determine the cleanliness level for which that cleanroom qualifies.

Set up the Test and Report Wizard

1. On the Counter Navigation screen, push **TEST WIZARD**.
2. In the Test and Report Wizard screen, complete the fields listed in this step. Select the applicable listed options.

Option	Description
Standard	Shows the list of applicable standards
Grade/Class	Shows a list of room classification
Occupancy	Shows a list of room states

3. Select the Sizes field.
 - a. Highlight a particle size.
 - b. Push **ADD** to add the size to the Considered Sizes list.
 - c. Add as many particle sizes as needed. To remove a particle size, select the size from the Considered Sizes list and push **REMOVE**.
4. Go to the next screen in the wizard.
 - a. Enter the room area in m². Push **ENTER** to confirm.
 - b. Enter the number of samples per location. Push **ENTER** to confirm.
5. Go to the next screen in the wizard.
 - a. Push **ADD AREA**, **ADD GROUP** or **ADD LOCATION**. Individual locations will show in the Locations list when the group, area or location is added.
 - b. Highlight a location. Use the **UP** or **DOWN** arrows to change the order of locations as needed. The order of the list will be the order of the test.
6. Go to the next screen. A confirmation of the selections will show.

Start sample measurement with the Wizard

1. Push **BEGIN SAMPLING**.
2. Obey the prompt and go to the first sample location. Push **OK**.
3. Put the isokinetic probe in position for the test. Push **RUN**.
4. Continue to obey the prompts and move to each location for the test. When all tests are complete, the counter will return to the wizard to configure the report.

Use existing data

The Test and Report Wizard can include existing data for the selected standard and location.

1. Select the Use Existing Data checkbox.
2. Enter the date range for the existing data.
3. Push **BEGIN SAMPLING**.

Report test results

- Push  to send the selected data to a USB drive. Refer to [Export data](#) on page 37.
- Push  to print the selected data. Refer to [How to use the Print Center](#) on page 35.

How to use the Print Center

About the Print Center

NOTICE

To avoid damage to the printer, do not operate the printer without paper. If the particle counter must be used without paper, be sure to set the print mode to "None".



The particle counter has a built-in printer. The Print Center screen is accessible from the:

- Counter Navigation screen
- Historical screen
- System Diagnostics screen
- Test/Report Wizard screen
- Area/Location Setup screen

On the Print Center screen the user can:

- Set automatic print functions
- Print buffer records or count averages

Note: Filtered data is printed from the Historical screen.

Print records manually

The buffer holds 5000 records maximum. The Print Center can print the entire buffer or the average of count cycles. To print records manually:

1. On the counter Navigation screen, push **PRINTER**.
2. On the Print Center screen, select the print option for the data.
 - Push **AVERAGE** to print the average of each size channel for the last sample measurement. Refer to [Figure 6](#).
 - Push **BUFFER** to print all of the records in the buffer. Refer to [Figure 7](#).
3. The data will begin to print.
 - To cancel the print job, push **CANCEL PRINT**.
 - To return to Counter Navigation, push **RETURN**.

Figure 6 Averages report





Figure 7 Buffer report (all buffer records)



Set automatic print functions

Note: If the sample period is very brief and the hold time is zero, some sample data may be skipped.

1. On the Counter Navigation screen, push **PRINTER**.
2. On the Print Center screen, select the Sample Print Mode field. Select an option for automatic printing.

Option	Description
None	No data will print automatically
Alarms	Prints results when a count alarm is exceeded
Cycles	Prints the results of the first count cycle and multiples of the programmed count cycle
All	Prints results after each count cycle is finished

3. Push **ENTER** to confirm.

Review historical buffer data



Records stored in the buffer are known as Historical data. These records can be accessed individually or sorted by location, date or time. The buffer can also be configured to collect partial data for aborted samples.

Set the data filter

The data filter can be set up to sort by location, date and time, either as individual parameters or in combination. To set the data filter, go to the Data Filter Setup screen.

1. Push  to go to the Data Filter Setup screen.
2. Select a filter option.

Option	Description
Filter by location	Check all of the locations to be included in the results. Push ALL to select all locations, or push NONE to deselect all locations.
Filter by date	Enter dates in the DATE FROM and DATE TO fields. Dates are in YYYY-MM-DD format.
Filter by time	Enter the needed times in the TIME FROM and TIME TO fields. Time is in HH:MM:SS format.

Clear the data buffer

Data that has been exported or is not needed can be deleted from the buffer.

1. On the Counter Navigation screen, push **SYSTEM**.
2. Select the **Options and Accessories** tab.
3. To erase all the data in the buffer, push **CLEAR BUFFER**.

Export data



Export data to maintain an electronic record for analysis and reporting.

1. On the Counter Navigation screen, push **EXPORT**.
2. Select **USB Flash Drive** or **Network server**.

Note: To export data to a network server, configure and enable the FTP function. Refer to [Configure and enable the FTP function](#) on page 39.

3. Select PDF, Comma Separated File, Tab Separated File, XML or PortAll.

Note: The PDF option shows only after the PDF option has been set to on. Refer to [Turn on the PDF option](#) on page 29.

4. If PDF was selected and the settings on the PDF tab under System Settings should be used, select the box to enable PDF Page and File Break Rules.
5. If needed, change the default file name and confirm.
6. If USB Flash Drive was selected, insert the USB drive into the USB host on the front of the instrument. If the USB Flash Drive is not inserted, a warning will show on the screen. The warning will change to a confirmation after the USB Flash Drive is inserted.
7. Push **EXPORT**. The status bar will show the progress.

Note: To stop the export, push .

About status values in exported data

The status value represents several elements of unit status. Environmental, count and concentration alarms can be indicated in the status value.

To determine the alarm conditions present in the report, subtract the largest possible value that represents a bit from the Status value in the exported data. Refer to [Table 4](#).

Table 4 Sample status bit mask definitions

Bit	Value	Definition	Bit	Value	Definition
0	1	Calibration	16	65536	Channel 3 concentration alarm
1	2	Flow	17	131072	Channel 4 concentration alarm
2	4	Temperature	18	262144	Channel 5 concentration alarm
3	8	Relative humidity	19	524288	Channel 6 concentration alarm
4	16	Air velocity	20	1048576	Channel 7 concentration alarm
5	32	System alarm	21	2097152	Channel 8 concentration alarm
6	64	Count alarm	22	4194304	Channel 9 concentration alarm
7	128	Concentration alarm	23	8388608	Channel 10 concentration alarm
8	256	Channel 1 count alarm	24	16777216	Channel 11 concentration alarm
9	512	Channel 2 count alarm	25	33554432	Channel 12 concentration alarm
10	1024	Channel 3 count alarm	26	67108864	Channel 7 count alarm
11	2048	Channel 4 count alarm	27	134217728	Channel 8 count alarm
12	4096	Channel 5 count alarm	28	268435456	Channel 9 count alarm
13	8192	Channel 6 count alarm	29	536870912	Channel 10 count alarm
14	16384	Channel 1 concentration alarm	30	1073741824	Channel 11 count alarm
15	32768	Channel 2 concentration alarm	31	2147483648	Channel 12 count alarm

Configure and enable the FTP function

With firmware V4.08, the particle counter can be configured to transmit data via Ethernet (wired or wireless when the wireless option is installed) to an FTP server. This particle counter can connect to servers that use FTP or FTPS (Explicit TLS/SSL).

1. For the user account that will be used in the particle counter, give *file* access on the FTP server: Read, Write and Delete.
2. For the user account that will be used in the particle counter, give *directory* access on the FTP server: Create, Delete, List and Add Subdirectories.

Note: The Create, Delete and Add Subdirectories rights are only necessary if the particle counter will be configured to make subdirectories. These rights are not necessary when the particle counter is configured to use the existing directories.

3. On the Counter Navigation screen, push **EXPORT**.
4. Select **Configure FTP**.
5. Select **Network Server Enabled**.
6. In the Host/IP Addr field, enter the host name or the IP address of the computer where the FTP server is installed.
7. In the Port field, enter the port number on which the FTP server listens.
8. In the Protocol field, select the protocol the FTP server uses (FTP or FTPS (TLS/SSL)).
9. Optional: In the Initial Folder field, enter the initial folder where files will be kept. Leave blank to save files to the root.
10. Select the **Logon** tab.
11. In the Authentication field, select Anonymous or Normal.
12. If Normal was selected, enter a user name and password to use for authentication with the FTP server.
13. Push **Test Connection**. When the connection is successful, "Test Connection Succeeded" shows.

Note: If the connection was not successful, an error message shows with the reason.

Data transfer to the OPC server

Only administrators, factory service personnel and operators with User Upload permission can initiate data upload to the OPC server.

There are three settings that control the data transfer to the OPC server: User initiated upload, user-initiated download and automatic download.

User-initiated data upload

Use this option to control when data is sent to the server.

1. Go to the Network screen.
2. Push  to initiate the data transfer.

User-initiated data download

User-initiated download is done from the OPC server. Use this option to verify the server connection and start the data download.

To start the data download, push **DOWNLOAD NOW**. All other actions are done by the server.

Automatic data download

Automatic download is configured on the OPC server. For applications that require alarm reasons, automatic download must be disabled.

Maintenance

⚠ WARNING

Multiple hazards. Do not disassemble the instrument for maintenance or service. If the internal components must be cleaned or repaired, contact the manufacturer.

⚠ CAUTION

Personal injury hazard. Only qualified personnel should conduct the tasks described in this section of the manual.

Clean the instrument exterior

The instrument exterior can be cleaned as needed. To avoid human exposure to potentially dangerous chemicals, make sure to clean the touchscreen immediately after contact with chemicals.

NOTICE

Do not leave visible moisture on the instrument or touchscreen. Moisture can penetrate the touchscreen and damage electronics inside.

1. Put the cap on the sample air intake nozzle.
2. Spray a mild cleaning solution on a soft cloth. Wipe the outside of the instrument carefully.
3. Use a soft, dry cloth to wipe the touchscreen surface. If needed, moisten the soft cloth with a mild cleaning solution.

Set the count to zero

Do this procedure after unexpectedly high particle counts. This procedure will verify that the particle counter works correctly and will remove residual particles.

1. Put the zero-count filter on the intake tube. Refer to [Figure 1](#) on page 10.
2. Turn on the unit and log in if needed.
3. Push **SAMPLE**.
4. Push **RUN**.
5. Repeat the process until the particle counts return to zero.

Update the instrument software

NOTICE

Do this procedure with only manufacturer-supplied files and directions.

1. Download the self-extracting *.zip file from the location provided by the manufacturer.
2. Extract the files to a compatible USB drive. Use only manufacturer-supplied USB drives for best results.
3. Remove AC power. Remove the batteries.
4. Plug the USB drive into the USB port.
5. Apply AC power to the instrument.
6. Select **Yes** to start the update. When the update is complete, the instrument will continue startup.
7. To verify the version number of the update, push **DIAGNOSTICS**. The version will show on the Diagnostics screen.

Charge the batteries in the particle counter

Batteries in the 3400 will begin to charge when the AC power cord is connected. A complete charge in the instrument takes approximately 10 hours. The battery is considered to be fully charged when they display shows the charge between 95% and 100%.

Prerequisite: Install the batteries in the instrument. Refer to [Install the batteries](#) on page 12.

NOTICE

Discard the used batteries according to local regulations or contact the manufacturer. Do not put exhausted batteries in the domestic waste.

1. Attach the power supply to the unit. Refer to [Figure 3](#) on page 12.
2. Connect the unit power supply to the external power through the AC power adapter.

The battery status light will show the level of power in the battery. Refer to [Table 5](#).

Table 5 Battery LED color indications

LED state	LED color	Battery status	Charge status
Flashing	Orange	Low power	Not charging
Flashing	Green	Low power	Charging
Solid	Green	Charged	Charging

Battery recharge intervals

[Table 6](#) shows the charge frequency that will increase battery life and increase the interval between battery calibrations.

Table 6 Suggested battery recharge interval

Frequency of use	Hours of sampling			
	0.5	1	2	6
Daily	Charge weekly	Charge weekly	Charge daily	Charge daily
Weekly	Charge monthly	Charge weekly	Charge weekly	Charge weekly
Monthly	Store on charge	Store on charge	Store on charge	Store on charge

When the particle counter is not in use, batteries will slowly discharge because of background processes on the instrument.

Calibrate the battery

The lithium ion Smart Battery will tolerate frequent partial discharges. After many partial discharges, the accuracy of the battery gauge is decreased.

To calibrate the battery charge gauge, set the Smart Charger to calibration mode during discharge.

Diagnostics and Troubleshooting



The Diagnostics screen shows information that may be needed for troubleshooting. [Table 7](#) shows an example of a failure notification on the Diagnostics screen.

Table 7 System Diagnostics screen example - Clock battery failure

Signal	Value	Status
Calibration	0.00 VDC	OFF
Flow	0.00 VDC	OFF
Clock battery	0.00 VDC	FAIL
Battery 1 (bottom)	16.42 VDC	PASS

Table 7 System Diagnostics screen example - Clock battery failure (continued)

Signal	Value	Status
Battery 2 (top)	16.44 VDC	PASS
Laser current	N/A	--

For troubleshooting that involves technical support from the manufacturer, the user may need to fax a system status printout to technical support.

1. On the Counter Navigation screen, push **DIAGNOSTICS**.
2. Go to the **Counter** tab to see information about the system, such as serial number, model and calibration date.
3. Go to the **Calibration** tab to see current calibration information.
4. On the Diagnostics screen, push **PRINT**.

The printout will show the serial number, date and time and other data about the system.

Factory settings



For troubleshooting that involves technical support from the manufacturer, the user may need to send a copy of the factory settings to technical support.

Note: *Factory settings cannot be changed by the user.*

To view the factory information including the calibration date:

1. Log on as an administrator.
2. On the Counter Navigation screen, push **FACTORY**. The factory settings show (e.g., calibration date, calibration due date and nominal flow).
3. Push  to print the factory settings.

Parts and accessories

⚠ WARNING

Multiple hazards. Use only the replacement parts and accessories specified by the manufacturer.

Parts for the 28.3 LPM counter (3413 and 3415)

Description	Quantity	Item number
Probe, isokinetic, aluminum, for 28.3 LPM	1	2087966-01
Probe, isokinetic, stainless steel, for 28.3 LPM	1	2087966-02
Filter, Zero Count for 28.3 LPM	1	2087939-01
Tubing, Hytrel®, 0.953 cm (0.375 in.) ID, 1.27 cm (0.5 in) OD	10 ft	960380

Parts for the 50 LPM counter (3423 and 3425)

Description	Quantity	Item number
Probe, isokinetic, aluminum, for 50 LPM	1	2088925-01
Probe, isokinetic, stainless steel, for 50 LPM	1	2088925-02

Parts for the 50 LPM counter (3423 and 3425) (continued)

Description	Quantity	Item number
Filter, Zero Count for 50 LPM	1	2087939-01
Tubing, Hytrel®, 0.953 cm (0.375 in.) ID, 1.27 cm (0.5 in) OD	10 ft	960380

Parts for the 100 LPM counter (3445)

Description	Quantity	Item number
Probe, isokinetic, aluminum, for 100 LPM	1	2088978-01
Probe, isokinetic, stainless steel, for 100 LPM	1	2088978-02
Filter, Zero Count for 100 LPM	1	2087939-02
Tubing, Hytrel®, 1.334 cm (0.525 in.) I. D.	10 ft	480-100-0041

3400 series parts

Description	Quantity	Item number
Battery, Lilon Smart	1	280-120-2024
Smart battery charger	1	280-300-5000
Brush, intake nozzle cleaning	1	995240
Paper, thermal, roll	1	460519
Power cord, 110 VAC	1	VP623501
Power cord, 220 VAC	1	VP6233500
Power supply, external, 100–240 VAC input, +24 VDC output	1	230-300-7052
WiFi antenna (for wireless versions only)	1	490-200-0001
Software, PortAll, (kit: CD, manual, serial adapter)	1	2084045-02
Software, PortAll, license	1	700011-21
Software, PortAll, license 21CFR	1	700011-22
Document, PortAll Software IQ/OQ	1	701169-01
Stylus for touchscreen interface	1	210-400-5171
Isokinetic probe stand (optional) (compatible with all Iso-kinetic probes), stainless steel	1	2089406-01
USB flash memory drive kit ¹	1	210-400-0128
Probe, air velocity, 0–200 fpm	1	826172
Probe, RH/T	1	2088928
Probe with beeper, filter scanning	1	2088000-01
Probe without beeper, filter scanning	1	2088000-02
Transportation case with foam inserts and wheels	1	710-200-0002
IQOQ validation document	1	2088789-01
Low pressure diffuser, 15–49 PSI, 1 CFM only	1	2088616-01
High pressure diffuser, 100 L/min and 50 L/min, 1/2-in. barb	1	2080732-11

3400 series parts (continued)

Description	Quantity	Item number
High pressure diffuser, 50 L/min (1.77 CFM), 3/8-in. barb	1	2080732-12
High pressure diffuser, 28.3 L/min (1.0 CFM), 3/8-in. barb	1	2080732-13
USB 2.0 high speed, 1 m (3.3 ft) cable	1	460-400-4798
USB to RS-232 adapter, DB-9 null modem	1	2088012-02
USB to RS-485 adapter	1	2088012-01

¹ Use only the manufacturer-supplied USB flash drives.

Spare parts kit (2087919-01)

Description	Quantity	Item number
Battery, Lilon Smart Battery	1	280-120-2024
Charger, external battery	1	280-300-5000
Thermal paper, roll	1	460519

**HACH COMPANY World Headquarters**

P.O. Box 389, Loveland, CO 80539-0389 U.S.A.

Tel. (970) 669-3050

(800) 227-4224 (U.S.A. only)

Fax (970) 669-2932

orders@hach.com

www.hach.com

HACH LANGE GMBH

Willstätterstraße 11

D-40549 Düsseldorf, Germany

Tel. +49 (0) 2 11 52 88-320

Fax +49 (0) 2 11 52 88-210

info@hach-lange.de

www.hach-lange.de

HACH LANGE Sàrl

6, route de Compois

1222 Vézenaz

SWITZERLAND

Tel. +41 22 594 6400

Fax +41 22 594 6499